

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD092920\
 Data File : VD066964.D
 Acq On : 30 Sep 2020 04:13
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 30 04:40:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 23 01:34:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.412	0.313	24.0#	82	0.00
3 P	Chloromethane	0.459	0.385	16.1	83	0.00
4 C	Vinyl Chloride	0.454	0.376	17.2#	79	0.00
5 T	Bromomethane	0.319	0.301	5.6	89	0.00
6 T	Chloroethane	0.282	0.253	10.3	85	0.00
7 T	Trichlorofluoromethane	0.932	0.830	10.9	85	0.00
8 T	Diethyl Ether	0.237	0.265	-11.8	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.540	0.493	8.7	87	0.00
10 T	Methyl Iodide	0.477	0.485	-1.7	88	0.00
11 T	Tert butyl alcohol	0.040	0.034	15.0	98	0.00
12 CM	1,1-Dichloroethene	0.490	0.477	2.7#	87	0.00
13 T	Acrolein	0.036	0.036	0.0	90	0.00
14 T	Allyl chloride	0.719	0.723	-0.6	88	0.00
15 T	Acrylonitrile	0.105	0.117	-11.4	96	0.00
16 T	Acetone	0.104	0.084	19.2	68	0.00
17 T	Carbon Disulfide	1.497	1.376	8.1	83	0.00
18 T	Methyl Acetate	0.258	0.270	-4.7	98	0.01
19 T	Methyl tert-butyl Ether	1.063	1.220	-14.8	97	0.00
20 T	Methylene Chloride	0.675	0.636	5.8	97	0.00
21 T	trans-1,2-Dichloroethene	0.567	0.606	-6.9	96	0.00
22 T	Diisopropyl ether	1.478	1.618	-9.5	93	0.00
23 T	Vinyl Acetate	0.727	0.837	-15.1	88	0.00
24 P	1,1-Dichloroethane	0.989	1.043	-5.5	94	0.00
25 T	2-Butanone	0.125	0.133	-6.4	91	0.00
26 T	2,2-Dichloropropane	0.906	0.765	15.6	78	0.00
27 T	cis-1,2-Dichloroethene	0.616	0.690	-12.0	97	0.00
28 T	Bromochloromethane	0.409	0.401	2.0	84	0.00
29 T	Tetrahydrofuran	0.080	0.089	-11.2	93	0.00
30 C	Chloroform	1.066	1.141	-7.0#	97	0.00
31 T	Cyclohexane	0.890	0.757	14.9	80	0.00
32 T	1,1,1-Trichloroethane	0.964	0.984	-2.1	92	0.00
33 S	1,2-Dichloroethane-d4	0.489	0.458	6.3	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.318	0.299	6.0	88	0.00
36 T	1,1-Dichloropropene	0.496	0.485	2.2	91	0.00
37 T	Ethyl Acetate	0.179	0.188	-5.0	90	0.00
38 T	Carbon Tetrachloride	0.542	0.526	3.0	91	0.00
39 T	Methylcyclohexane	0.570	0.540	5.3	85	0.00
40 TM	Benzene	1.382	1.450	-4.9	96	0.00
41 T	Methacrylonitrile	0.110	0.131	-19.1	108	0.00
42 TM	1,2-Dichloroethane	0.409	0.415	-1.5	95	0.00
43 T	Isopropyl Acetate	0.346	0.362	-4.6	92	0.00
44 TM	Trichloroethene	0.394	0.413	-4.8	98	0.00
45 C	1,2-Dichloropropane	0.348	0.359	-3.2#	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD092920\
 Data File : VD066964.D
 Acq On : 30 Sep 2020 04:13
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 30 04:40:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 23 01:34:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.189	0.204	-7.9	95	0.00
47 T	Bromodichloromethane	0.508	0.537	-5.7	98	0.00
48 T	Methyl methacrylate	0.170	0.177	-4.1	93	0.00
49 T	1,4-Dioxane	0.002	0.003	-50.0#	101	0.00
50 S	Toluene-d8	1.084	1.044	3.7	87	0.00
51 T	4-Methyl-2-Pentanone	0.175	0.193	-10.3	94	0.00
52 CM	Toluene	0.890	0.961	-8.0#	97	0.00
53 T	t-1,3-Dichloropropene	0.432	0.471	-9.0	95	0.00
54 T	cis-1,3-Dichloropropene	0.527	0.561	-6.5	94	0.00
55 T	1,1,2-Trichloroethane	0.260	0.283	-8.8	101	0.00
56 T	Ethyl methacrylate	0.278	0.327	-17.6	99	0.00
57 T	1,3-Dichloropropane	0.429	0.473	-10.3	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.132	0.128	3.0	92	0.00
59 T	2-Hexanone	0.118	0.129	-9.3	91	0.00
60 T	Dibromochloromethane	0.339	0.379	-11.8	102	0.00
61 T	1,2-Dibromoethane	0.242	0.269	-11.2	100	0.00
62 S	4-Bromofluorobenzene	0.371	0.369	0.5	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.360	0.387	-7.5	105	0.00
65 PM	Chlorobenzene	1.041	1.120	-7.6	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.432	-9.6	104	0.00
67 C	Ethyl Benzene	1.828	1.948	-6.6#	98	0.00
68 T	m/p-Xylenes	0.692	0.751	-8.5	99	0.00
69 T	o-Xylene	0.626	0.700	-11.8	101	0.00
70 T	Styrene	1.087	1.226	-12.8	101	0.00
71 P	Bromoform	0.199	0.226	-13.6	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.612	3.940	-9.1	99	0.00
74 T	N-amyl acetate	0.705	0.749	-6.2	93	0.00
75 P	1,1,2,2-Tetrachloroethane	0.623	0.664	-6.6	100	0.00
76 T	1,2,3-Trichloropropane	0.414	0.434	-4.8	100	0.00
77 T	Bromobenzene	0.845	0.939	-11.1	103	0.00
78 T	n-propylbenzene	4.345	4.569	-5.2	97	0.00
79 T	2-Chlorotoluene	2.476	2.648	-6.9	101	0.00
80 T	1,3,5-Trimethylbenzene	3.065	3.351	-9.3	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.166	0.185	-11.4	98	0.00
82 T	4-Chlorotoluene	2.610	2.792	-7.0	100	0.00
83 T	tert-Butylbenzene	2.518	2.870	-14.0	105	0.00
84 T	1,2,4-Trimethylbenzene	3.039	3.401	-11.9	103	0.00
85 T	sec-Butylbenzene	3.622	3.811	-5.2	98	0.00
86 T	p-Isopropyltoluene	3.264	3.598	-10.2	101	0.00
87 T	1,3-Dichlorobenzene	1.669	1.788	-7.1	102	0.00
88 T	1,4-Dichlorobenzene	1.658	1.783	-7.5	105	0.00
89 T	n-Butylbenzene	3.103	3.200	-3.1	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD092920\
Data File : VD066964.D
Acq On : 30 Sep 2020 04:13
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Sep 30 04:40:18 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 23 01:34:05 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.663	0.684	-3.2	100	0.00
91 T	1,2-Dichlorobenzene	1.428	1.574	-10.2	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.100	0.108	-8.0	102	0.00
93 T	1,2,4-Trichlorobenzene	0.899	1.026	-14.1	107	0.00
94 T	Hexachlorobutadiene	0.539	0.577	-7.1	101	0.00
95 T	Naphthalene	1.526	1.893	-24.0#	106	0.00
96 T	1,2,3-Trichlorobenzene	0.776	0.911	-17.4	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6