

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092921\
 Data File : VD070521.D
 Acq On : 29 Sep 2021 14:23
 Operator : VA/SY
 Sample : M3938-03
 Misc : 6.72G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SUB-SLAB-05(6-2.6)

Quant Time: Sep 30 00:45:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	464931	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1002170	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1010877	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	460381	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	338665	56.730	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.460%
35) Dibromofluoromethane	7.914	113	351331	49.250	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.500%
50) Toluene-d8	10.338	98	1310729	50.607	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.220%
62) 4-Bromofluorobenzene	12.620	95	515986	57.004	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	114.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092921\
Data File : VD070521.D
Acq On : 29 Sep 2021 14:23
Operator : VA/SY
Sample : M3938-03
Misc : 6.72G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SUB-SLAB-05(6-2.6)

Quant Time: Sep 30 00:45:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

