

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092921\
 Data File : VD070522.D
 Acq On : 29 Sep 2021 14:50
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Sep 30 00:45:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	380806	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	725192	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	688928	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	325317	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	244213	49.945	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.900%
35) Dibromofluoromethane	7.914	113	250537	48.535	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.060%
50) Toluene-d8	10.338	98	869937	46.416	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.840%
62) 4-Bromofluorobenzene	12.620	95	338112	51.619	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092921\
Data File : VD070522.D
Acq On : 29 Sep 2021 14:50
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Sep 30 00:45:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

