

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092921\
 Data File : VD070524.D
 Acq On : 29 Sep 2021 15:47
 Operator : VA/SY
 Sample : M3975-02
 Misc : 6.89G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-837-R-SO-4-4.5-092821

Quant Time: Sep 30 00:45:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	456464	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	953219	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	876208	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	358604	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	317364	54.148	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 108.300%			
35) Dibromofluoromethane	7.914	113	334310	49.271	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 98.540%			
50) Toluene-d8	10.332	98	1214350	49.293	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 98.580%			
62) 4-Bromofluorobenzene	12.626	95	419691	48.746	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 97.500%			
Target Compounds					Qvalue	
16) Acetone	4.156	43	70721	67.584	ug/l	97
25) 2-Butanone	7.214	43	18532	13.304	ug/l	95
95) Naphthalene	15.385	128	20592	1.969	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092921\
Data File : VD070524.D
Acq On : 29 Sep 2021 15:47
Operator : VA/SY
Sample : M3975-02
Misc : 6.89G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-837-R-SO-4-4.5-092821

Quant Time: Sep 30 00:45:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

