

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092921\
 Data File : VD070527.D
 Acq On : 29 Sep 2021 17:11
 Operator : VA/SY
 Sample : M3975-05
 Misc : 8.44G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-841-R-SO-2-2.5-092821

Quant Time: Sep 30 00:46:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	159949	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	322298	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	334801	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	154663	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	151705	73.866	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 147.740%	
35) Dibromofluoromethane	7.914	113	128998	56.229	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 112.460%	
50) Toluene-d8	10.338	98	423596	50.855	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 101.700%	
62) 4-Bromofluorobenzene	12.626	95	155495	53.415	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 106.840%	
Target Compounds						
16) Acetone	4.162	43	20708	56.476	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092921\
Data File : VD070527.D
Acq On : 29 Sep 2021 17:11
Operator : VA/SY
Sample : M3975-05
Misc : 8.44G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-841-R-SO-2-2.5-092821

Quant Time: Sep 30 00:46:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

