

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD093021\
 Data File : VD070539.D
 Acq On : 30 Sep 2021 14:24
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Oct 01 02:04:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	269784	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	525142	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	491972	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	207548	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	180738	52.175	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 104.340%	
35) Dibromofluoromethane	7.914	113	185875	49.725	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.460%	
50) Toluene-d8	10.338	98	628861	46.336	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 92.680%	
62) 4-Bromofluorobenzene	12.626	95	190096	40.078	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 80.160%	
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	7.214	96	19649	4.892	ug/l	88
44) Trichloroethene	9.108	130	42249	10.747	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD093021\
Data File : VD070539.D
Acq On : 30 Sep 2021 14:24
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Oct 01 02:04:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

