

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD093021\
 Data File : VD070532.D
 Acq On : 30 Sep 2021 10:23
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 01 02:02:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.574	0.479	16.6	87	0.00
3 P	Chloromethane	0.771	0.595	22.8	80	0.00
4 C	Vinyl Chloride	0.787	0.737	6.4#	93	0.00
5 T	Bromomethane	0.580	0.465	19.8	89	0.00
6 T	Chloroethane	0.541	0.483	10.7	94	0.00
7 T	Trichlorofluoromethane	1.041	1.007	3.3	90	0.00
8 T	Diethyl Ether	0.288	0.269	6.6	80	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.646	0.637	1.4	91	0.00
10 T	Methyl Iodide	0.564	0.597	-5.9	81	0.00
11 T	Tert butyl alcohol	0.073	0.027	63.0#	47#	0.00
12 CM	1,1-Dichloroethene	0.582	0.572	1.7#	88	0.00
13 T	Acrolein	0.032	0.025	21.9	72	0.00
14 T	Allyl chloride	0.855	0.815	4.7	87	0.00
15 T	Acrylonitrile	0.137	0.123	10.2	82	0.00
16 T	Acetone	0.115	0.128	-11.3	107	0.00
17 T	Carbon Disulfide	2.148	1.803	16.1	78	0.00
18 T	Methyl Acetate	0.344	0.284	17.4	83	0.00
19 T	Methyl tert-butyl Ether	1.258	1.214	3.5	81	0.00
20 T	Methylene Chloride	0.913	0.666	27.1#	79	0.00
21 T	trans-1,2-Dichloroethene	0.693	0.669	3.5	86	0.00
22 T	Diisopropyl ether	1.831	1.788	2.3	86	0.00
23 T	Vinyl Acetate	0.848	0.809	4.6	84	0.00
24 P	1,1-Dichloroethane	1.295	1.190	8.1	86	0.00
25 T	2-Butanone	0.153	0.140	8.5	87	0.00
26 T	2,2-Dichloropropane	1.047	1.055	-0.8	95	0.00
27 T	cis-1,2-Dichloroethene	0.744	0.721	3.1	85	0.00
28 T	Bromochloromethane	0.514	0.490	4.7	89	0.00
29 T	Tetrahydrofuran	0.098	0.089	9.2	83	0.00
30 C	Chloroform	1.328	1.248	6.0#	88	0.00
31 T	Cyclohexane	1.106	1.058	4.3	89	0.00
32 T	1,1,1-Trichloroethane	1.118	1.106	1.1	92	0.00
33 S	1,2-Dichloroethane-d4	0.642	0.530	17.4	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.356	0.333	6.5	82	0.00
36 T	1,1-Dichloropropene	0.534	0.555	-3.9	89	0.00
37 T	Ethyl Acetate	0.200	0.184	8.0	85	0.00
38 T	Carbon Tetrachloride	0.529	0.567	-7.2	92	0.00
39 T	Methylcyclohexane	0.597	0.651	-9.0	87	0.00
40 TM	Benzene	1.566	1.586	-1.3	86	0.00
41 T	Methacrylonitrile	0.105	0.102	2.9	90	0.00
42 TM	1,2-Dichloroethane	0.419	0.390	6.9	84	0.00
43 T	Isopropyl Acetate	0.370	0.331	10.5	81	0.00
44 TM	Trichloroethene	0.374	0.378	-1.1	84	0.00
45 C	1,2-Dichloropropane	0.411	0.401	2.4#	86	0.00
46 T	Dibromomethane	0.212	0.196	7.5	82	0.00
47 T	Bromodichloromethane	0.540	0.519	3.9	84	0.00
48 T	Methyl methacrylate	0.171	0.164	4.1	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD093021\
 Data File : VD070532.D
 Acq On : 30 Sep 2021 10:23
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 01 02:02:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	81	0.00
50 S	Toluene-d8	1.292	1.284	0.6	83	0.00
51 T	4-Methyl-2-Pentanone	0.191	0.180	5.8	82	0.00
52 CM	Toluene	0.946	1.006	-6.3#	86	0.00
53 T	t-1,3-Dichloropropene	0.464	0.472	-1.7	86	0.00
54 T	cis-1,3-Dichloropropene	0.565	0.581	-2.8	90	0.00
55 T	1,1,2-Trichloroethane	0.291	0.276	5.2	81	0.00
56 T	Ethyl methacrylate	0.318	0.327	-2.8	81	0.00
57 T	1,3-Dichloropropane	0.496	0.477	3.8	82	0.00
58 T	2-Chloroethyl Vinyl ether	0.153	0.161	-5.2	88	0.00
59 T	2-Hexanone	0.128	0.134	-4.7	89	0.00
60 T	Dibromochloromethane	0.339	0.327	3.5	80	0.00
61 T	1,2-Dibromoethane	0.268	0.261	2.6	83	0.00
62 S	4-Bromofluorobenzene	0.452	0.431	4.6	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	81	0.00
64 T	Tetrachloroethene	0.304	0.335	-10.2	87	0.00
65 PM	Chlorobenzene	1.036	1.086	-4.8	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.379	0.401	-5.8	86	0.00
67 C	Ethyl Benzene	1.786	2.035	-13.9#	88	0.00
68 T	m/p-Xylenes	0.690	0.796	-15.4	87	0.00
69 T	o-Xylene	0.618	0.724	-17.2	87	0.00
70 T	Styrene	1.080	1.267	-17.3	85	0.00
71 P	Bromoform	0.187	0.185	1.1	79	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	3.506	4.147	-18.3	90	0.00
74 T	N-amyl acetate	0.764	0.746	2.4	83	0.00
75 P	1,1,2,2-Tetrachloroethane	0.751	0.692	7.9	80	0.00
76 T	1,2,3-Trichloropropane	0.506	0.467	7.7	78	0.00
77 T	Bromobenzene	0.809	0.830	-2.6	83	0.00
78 T	n-propylbenzene	4.551	5.317	-16.8	90	0.00
79 T	2-Chlorotoluene	2.653	2.962	-11.6	88	0.00
80 T	1,3,5-Trimethylbenzene	3.026	3.544	-17.1	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.202	0.204	-1.0	84	0.00
82 T	4-Chlorotoluene	2.799	3.114	-11.3	87	0.00
83 T	tert-Butylbenzene	2.427	2.934	-20.9	91	0.00
84 T	1,2,4-Trimethylbenzene	2.980	3.493	-17.2	88	0.00
85 T	sec-Butylbenzene	3.858	4.614	-19.6	92	0.00
86 T	p-Isopropyltoluene	3.100	3.770	-21.6	91	0.00
87 T	1,3-Dichlorobenzene	1.684	1.799	-6.8	85	0.00
88 T	1,4-Dichlorobenzene	1.691	1.721	-1.8	84	0.00
89 T	n-Butylbenzene	3.091	3.627	-17.3	91	0.00
90 T	Hexachloroethane	0.710	0.737	-3.8	89	0.00
91 T	1,2-Dichlorobenzene	1.486	1.497	-0.7	82	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.112	0.101	9.8	79	0.00
93 T	1,2,4-Trichlorobenzene	0.802	0.839	-4.6	84	0.00
94 T	Hexachlorobutadiene	0.492	0.529	-7.5	94	0.00
95 T	Naphthalene	1.458	1.527	-4.7	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD093021\
Data File : VD070532.D
Acq On : 30 Sep 2021 10:23
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Oct 01 02:02:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.708	0.727	-2.7	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6