

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD100118\
 Data File : VD060099.D
 Acq On : 1 Oct 2018 16:24
 Operator : VA/AP
 Sample : VSTDIC005
 Misc : 5.00g/5ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC005

Quant Time: Oct 02 02:09:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:08:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.39	168	1598506	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	7.42	114	2257213	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.27	117	1805814	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.36	152	952202	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.76	65	63158	5.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	10.14%	
35) Dibromofluoromethane	6.30	113	86501	4.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	9.80%	
50) Toluene-d8	9.44	98	241825	4.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	9.78%	
62) 4-Bromofluorobenzene	12.42	95	91865	4.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	9.84%	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.57	85	80979	4.900 ug/l	94
3) Chloromethane	1.74	50	73881	4.995 ug/l #	87
4) Vinyl Chloride	1.83	62	61183	5.153 ug/l	96
5) Bromomethane	2.12	94	10180	7.389 ug/l #	78
6) Chloroethane	2.23	64	16893	6.399 ug/l	100
7) Trichlorofluoromethane	2.48	101	72122	5.521 ug/l	94
8) Diethyl Ether	2.80	74	14287	5.139 ug/l	94
9) 1,1,2-Trichlorotrifluoroet	3.07	101	46803	5.705 ug/l	96
10) Methyl Iodide	3.22	142	38202	4.364 ug/l #	86
11) Tert butyl alcohol	3.95	59	22805	23.065 ug/l	97
12) 1,1-Dichloroethene	3.05	96	36766	5.660 ug/l	93
13) Acrolein	2.96	56	16478	26.976 ug/l	96
14) Allyl chloride	3.51	41	71716	5.248 ug/l	97
15) Acrylonitrile	4.05	53	95369	25.449 ug/l	95
16) Acetone	3.14	43	64054	27.114 ug/l	97
17) Carbon Disulfide	3.29	76	117413	5.258 ug/l	98
18) Methyl Acetate	3.53	43	25584	5.259 ug/l	97
19) Methyl tert-butyl Ether	4.09	73	151846	5.065 ug/l	99
20) Methylene Chloride	3.69	84	149690	7.430 ug/l	97
21) trans-1,2-Dichloroethene	4.06	96	88946	5.211 ug/l	95
22) Diisopropyl ether	4.83	45	326816	5.268 ug/l	99
23) Vinyl Acetate	4.77	43	913237	25.937 ug/l	97
24) 1,1-Dichloroethane	4.73	63	141493	5.051 ug/l	99
25) 2-Butanone	5.60	43	138359	21.709 ug/l	96
26) 2,2-Dichloropropane	5.56	77	114179	5.267 ug/l	100
27) cis-1,2-Dichloroethene	5.57	96	94207	5.280 ug/l	96
28) Bromochloromethane	5.90	49	71013	5.216 ug/l	99
29) Tetrahydrofuran	5.92	42	86114	25.194 ug/l	97
30) Chloroform	6.08	83	146404	5.175 ug/l	98
31) Cyclohexane	6.35	56	157752	5.818 ug/l #	93
32) 1,1,1-Trichloroethane	6.27	97	119063	5.092 ug/l	98
36) 1,1-Dichloropropene	6.49	75	116179	5.261 ug/l	98
37) Ethyl Acetate	5.68	43	74077	4.858 ug/l	98
38) Carbon Tetrachloride	6.47	117	101705	4.959 ug/l	96

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD100118\
 Data File : VD060099.D
 Acq On : 1 Oct 2018 16:24
 Operator : VA/AP
 Sample : VSTDIC005
 Misc : 5.00g/5ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC005

Quant Time: Oct 02 02:09:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:08:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	8.02	83	126848	5.062	ug/l	99
40) Benzene	6.77	78	296402	5.259	ug/l	99
41) Methacrylonitrile	5.93	41	85076	10.486	ug/l #	86
42) 1,2-Dichloroethane	6.87	62	82647	4.889	ug/l	99
43) Isopropyl Acetate	8.32	43	92324	4.569	ug/l #	100
44) Trichloroethene	7.71	130	88702	5.067	ug/l	96
45) 1,2-Dichloropropane	8.08	63	74351	4.928	ug/l	98
46) Dibromomethane	8.20	93	48160	4.911	ug/l	96
47) Bromodichloromethane	8.47	83	101276	4.778	ug/l	100
48) Methyl methacrylate	8.23	41	52777	4.685	ug/l	95
49) 1,4-Dioxane	8.21	88	10332	92.168	ug/l	90
51) 4-Methyl-2-Pentanone	9.33	43	306288	22.096	ug/l	98
52) Toluene	9.54	92	186233	5.120	ug/l	95
53) t-1,3-Dichloropropene	9.90	75	90587	4.576	ug/l	95
54) cis-1,3-Dichloropropene	9.08	75	112344	4.756	ug/l	94
55) 1,1,2-Trichloroethane	10.16	97	60958	4.832	ug/l	96
56) Ethyl methacrylate	10.02	69	61854	4.448	ug/l	97
57) 1,3-Dichloropropane	10.37	76	91142	4.905	ug/l	96
58) 2-Chloroethyl Vinyl ether	8.91	63	194032	28.294	ug/l	95
59) 2-Hexanone	10.47	43	214434	21.437	ug/l	100
60) Dibromochloromethane	10.63	129	68462	4.498	ug/l	95
61) 1,2-Dibromoethane	10.75	107	61159	4.697	ug/l	99
64) Tetrachloroethene	10.23	164	81400	5.175	ug/l	99
65) Chlorobenzene	11.30	112	202494	5.273	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.40	131	66005	5.109	ug/l	97
67) Ethyl Benzene	11.41	91	349227	5.664	ug/l	99
68) m/p-Xylenes	11.55	106	252561	11.122	ug/l	89
69) o-Xylene	11.93	106	122586	5.413	ug/l	98
70) Styrene	11.95	104	202422	5.323	ug/l	98
71) Bromoform	12.10	173	47362	4.283	ug/l	96
73) Isopropylbenzene	12.27	105	336457	5.438	ug/l	100
74) N-amyl acetate	12.12	43	118321	4.859	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.56	83	71475	5.096	ug/l	98
76) 1,2,3-Trichloropropane	12.59	75	68200	4.765	ug/l	92
77) Bromobenzene	12.53	156	97342	5.430	ug/l	98
78) n-propylbenzene	12.63	91	435918	5.620	ug/l	98
79) 2-Chlorotoluene	12.69	91	231837	5.305	ug/l	98
80) 1,3,5-Trimethylbenzene	12.78	105	269106	5.677	ug/l	100
81) trans-1,4-Dichloro-2-buten	12.33	75	17901	4.151	ug/l	94
82) 4-Chlorotoluene	12.80	91	279993	5.827	ug/l	100
83) tert-Butylbenzene	13.04	119	291291	5.293	ug/l	99
84) 1,2,4-Trimethylbenzene	13.09	105	275179	5.423	ug/l	98
85) sec-Butylbenzene	13.22	105	357723	5.508	ug/l	98
86) p-Isopropyltoluene	13.33	119	291827	5.552	ug/l	99
87) 1,3-Dichlorobenzene	13.30	146	169772	5.456	ug/l	99
88) 1,4-Dichlorobenzene	13.38	146	172340	5.465	ug/l	97
89) n-Butylbenzene	13.64	91	302060	5.900	ug/l	100
90) Hexachloroethane	13.85	117	57682	4.438	ug/l	95
91) 1,2-Dichlorobenzene	13.65	146	152262	5.929	ug/l	97
92) 1,2-Dibromo-3-Chloropropan	14.22	75	8855	4.133	ug/l	87

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD100118\
Data File : VD060099.D
Acq On : 1 Oct 2018 16:24
Operator : VA/AP
Sample : VSTDICC005
Misc : 5.00g/5ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC005

Quant Time: Oct 02 02:09:19 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100118S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:08:01 2018
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.78	180	115645	5.215	ug/l	98
94) Hexachlorobutadiene	14.87	225	75044	5.108	ug/l	96
95) Naphthalene	14.96	128	164874	4.821	ug/l	100
96) 1,2,3-Trichlorobenzene	15.11	180	100398	5.169	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD100118\
Data File : VD060099.D
Acq On : 1 Oct 2018 16:24
Operator : VA/AP
Sample : VSTDIC005
Misc : 5.00g/5ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC005

Quant Time: Oct 02 02:09:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D100118S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:08:01 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 6.39 min Scan# 494

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :


VSTDIC005

Tgt Ion:168 Resp: 1598506

Ion Ratio Lower Upper

168 100

99 45.3 35.4 53.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.95 (98.65 to 99.65): VDC

6.39



600000

400000

200000

0

Time--> 6.20 6.30 6.40 6.50 6.60



#2

Dichlorodifluoromethane

Concen: 4.900 ug/l

RT: 1.57 min Scan# 5

Delta R.T. 0.01 min

Lab File: VD060099.D


Acq: 1 Oct 2018 16:24

Tgt Ion: 85 Resp: 80979

Ion Ratio Lower Upper

85 100

87 28.1 15.9 47.6



Abundance Ion 84.95 (84.65 to 85.65): VDC

Ion 86.90 (86.60 to 87.60): VDC

1.57

50000

40000


30000


20000

10000

0

Time--> 1.50 1.60 1.70





#3

Chloromethane

Concen: 4.995 ug/l

RT: 1.74 min Scan# 22

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :


VSTDIC005

Tgt Ion: 50 Resp: 73881

Ion Ratio Lower Upper




50 100

52 39.0 25.3 37.9#



Abundance Ion 49.90 (49.60 to 50.60): VDC

Ion 51.90 (51.60 to 52.60): VDC



#4

Vinyl Chloride

Concen: 5.153 ug/l

RT: 1.83 min Scan# 31

Delta R.T. 0.00 min

Lab File: VD060099.D


Acq: 1 Oct 2018 16:24

Tgt Ion: 62 Resp: 61183

Ion Ratio Lower Upper



62 100


64 33.5 25.2 37.8



Abundance Ion 61.95 (61.65 to 62.65): VDC

Ion 63.95 (63.65 to 64.65): VDC





#5

Bromomethane

Concen: 7.389 ug/l

RT: 2.12 min Scan# 61

Delta R.T. 0.01 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :


VSTDIC005

Tgt Ion: 94 Resp: 10180

Ion Ratio Lower Upper




94 100

96 74.7 77.4 116.2#



Abundance Ion 93.90 (93.60 to 94.60): VDC

Ion 95.90 (95.60 to 96.60): VDC



#6

Chloroethane

Concen: 6.399 ug/l

RT: 2.23 min Scan# 72

Delta R.T. 0.00 min

Lab File: VD060099.D


Acq: 1 Oct 2018 16:24

Tgt Ion: 64 Resp: 16893

Ion Ratio Lower Upper



64 100


66 30.6 24.6 37.0



Abundance Ion 63.95 (63.65 to 64.65): VDC

Ion 65.90 (65.60 to 66.60): VDC






#7




Trichlorofluoromethane
 Concen: 5.521 ug/l
 RT: 2.48 min Scan# 97
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC005

Tgt Ion: 101 Resp: 72122
 Ion Ratio Lower Upper
 101 100
 103 69.2 51.5 77.3




Abundance Ion 100.85 (100.55 to 101.55): V
 Ion 102.85 (102.55 to 103.55): V





#8


Diethyl Ether
 Concen: 5.139 ug/l
 RT: 2.80 min Scan# 130
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Tgt Ion: 74 Resp: 14287
 Ion Ratio Lower Upper
 74 100
 45 153.3 80.3 241.0



Abundance Ion 74.00 (73.70 to 74.70): VDC
 Ion 45.05 (44.75 to 45.75): VDC





#9

1,1,2-Trichlorotrifluoroethane

Concen: 5.705 ug/l

RT: 3.07 min Scan# 157

Delta R.T. 0.01 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC005


Tgt Ion:101 Resp: 46803

Ion Ratio Lower Upper

101 100

85 37.0 33.4 50.0




151 61.4 48.0 72.0



Abundance Ion 100.85 (100.55 to 101.55): V

Ion 84.95 (84.65 to 85.65): V

Ion 150.90 (150.60 to 151.60): V



#10

Methyl Iodide

Concen: 4.364 ug/l

RT: 3.22 min Scan# 172

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24


Tgt Ion:142 Resp: 38202

Ion Ratio Lower Upper

142 100

127 64.8 42.0 63.0#



141 13.5 11.4 17.0




Abundance Ion 141.85 (141.55 to 142.55): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.85 (140.55 to 141.55): V









#11

Tert butyl alcohol
 Concen: 23.065 ug/l
 RT: 3.95 min Scan# 246
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC005




Tgt Ion: 59 Resp: 22805
 Ion Ratio Lower Upper
 59 100
 57 10.9 7.8 11.6




#12

1,1-Dichloroethene
 Concen: 5.660 ug/l
 RT: 3.05 min Scan# 155
 Delta R.T. 0.01 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Tgt Ion: 96 Resp: 36766
 Ion Ratio Lower Upper
 96 100
 61 178.4 153.0 229.4
 98 65.8 51.1 76.7





#13

Acrolein

Concen: 26.976 ug/l

RT: 2.96 min Scan# 146

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :





VSTDIC005

Tgt Ion: 56 Resp: 16478

Ion Ratio Lower Upper

56 100

55 75.4 57.4 86.2



#14

Allyl chloride

Concen: 5.248 ug/l

RT: 3.51 min Scan# 202

Delta R.T. 0.01 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24




Tgt Ion: 41 Resp: 71716


Ion Ratio Lower Upper

41 100

39 78.6 60.6 90.8

76 25.9 21.3 31.9





#15

Acrylonitrile

Concen: 25.449 ug/l

RT: 4.05 min Scan# 257

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC005





Tgt Ion: 53 Resp: 95369

Ion Ratio Lower Upper

53 100

52 88.8 67.5 101.3

51 39.4 29.3 43.9



#16

Acetone

Concen: 27.114 ug/l

RT: 3.14 min Scan# 164

Delta R.T. 0.00 min

Lab File: VD060099.D




Acq: 1 Oct 2018 16:24


Tgt Ion: 43 Resp: 64054

Ion Ratio Lower Upper

43 100

58 20.7 17.7 26.5









#17
Carbon Disulfide
Concen: 5.258 ug/l
RT: 3.29 min Scan# 179
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC005


Tgt Ion: 76 Resp: 117413
Ion Ratio Lower Upper
76 100
78 8.5 7.4 11.2



#18
Methyl Acetate
Concen: 5.259 ug/l
RT: 3.53 min Scan# 204
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Tgt Ion: 43 Resp: 25584
Ion Ratio Lower Upper
43 100
74 14.5 12.6 19.0









#19
Methyl tert-butyl Ether
Concen: 5.065 ug/l
RT: 4.09 min Scan# 261
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC005


Tgt Ion: 73 Resp: 151846
Ion Ratio Lower Upper
73 100
57 23.7 18.7 28.1



#20
Methylene Chloride
Concen: 7.430 ug/l
RT: 3.69 min Scan# 220
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Tgt Ion: 84 Resp: 149690
Ion Ratio Lower Upper
84 100
49 146.9 114.6 172.0
51 41.8 33.8 50.8
86 69.1 53.3 79.9





#21

trans-1,2-Dichloroethene

Concen: 5.211 ug/l

RT: 4.06 min Scan# 258

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC005


Tgt Ion: 96 Resp: 88946

Ion Ratio Lower Upper




96 100

61 135.0 103.3 154.9

98 59.8 50.4 75.6



Abundance Ion 95.90 (95.60 to 96.60): VDC
Ion 60.95 (60.65 to 61.65): VDC
Ion 97.85 (97.55 to 98.55): VDC



#22

Diisopropyl ether

Concen: 5.268 ug/l

RT: 4.83 min Scan# 336

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Tgt Ion: 45 Resp: 326816


Ion Ratio Lower Upper

45 100



43 56.3 44.8 67.2


87 18.6 13.9 20.9

59 9.7 7.1 10.7



Abundance Ion 45.05 (44.75 to 45.75): VDC
Ion 43.05 (42.75 to 43.75): VDC
Ion 87.00 (86.70 to 87.70): VDC
Ion 58.95 (58.65 to 59.65): VDC





#23

Vinyl Acetate

Concen: 25.937 ug/l

RT: 4.77 min Scan# 330

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :


VSTDIC005

Tgt Ion: 43 Resp: 913237

Ion Ratio Lower Upper




43 100

86 5.9 5.6 8.4



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 86.00 (85.70 to 86.70): VDC



#24

1,1-Dichloroethane

Concen: 5.051 ug/l

RT: 4.73 min Scan# 326

Delta R.T. 0.01 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24


Tgt Ion: 63 Resp: 141493

Ion Ratio Lower Upper

63 100

98 5.2 2.8 8.3



100 4.2 1.8 5.3




Abundance Ion 62.95 (62.65 to 63.65): VDC

Ion 97.95 (97.65 to 98.65): VDC

Ion 99.95 (99.65 to 100.65): VDC





#25

2-Butanone

Concen: 21.709 ug/l

RT: 5.60 min Scan# 414

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :


VSTDIC005

Tgt Ion: 43 Resp: 138359




Ion Ratio Lower Upper

43 100

72 16.5 14.5 21.7



Abundance Ion 42.95 (42.65 to 43.65): VDC
Ion 72.00 (71.70 to 72.70): VDC



#26

2,2-Dichloropropane

Concen: 5.267 ug/l

RT: 5.56 min Scan# 410

Delta R.T. 0.00 min

Lab File: VD060099.D


Acq: 1 Oct 2018 16:24

Tgt Ion: 77 Resp: 114179



Ion Ratio Lower Upper


77 100

97 26.9 13.3 39.9



Abundance Ion 76.95 (76.65 to 77.65): VDC
Ion 96.85 (96.55 to 97.55): VDC









#27

cis-1,2-Dichloroethene
 Concen: 5.280 ug/l
 RT: 5.57 min Scan# 411
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC005




Tgt Ion: 96 Resp: 94207
 Ion Ratio Lower Upper
 96 100
 61 127.3 0.0 251.8
 98 70.5 0.0 125.2




#28

Bromochloromethane
 Concen: 5.216 ug/l
 RT: 5.90 min Scan# 445
 Delta R.T. 0.01 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Tgt Ion: 49 Resp: 71013
 Ion Ratio Lower Upper
 49 100
 129 1.8 0.0 3.6
 130 68.4 53.9 80.9









#29
Tetrahydrofuran
Concen: 25.194 ug/l
RT: 5.92 min Scan# 447
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC005

Tgt Ion: 42 Resp: 86114
Ion Ratio Lower Upper
42 100
72 32.9 26.9 40.3
71 31.2 26.7 40.1




Abundance Ion 42.00 (41.70 to 42.70): VDC
Ion 72.00 (71.70 to 72.70): VDC
Ion 71.00 (70.70 to 71.70): VDC






#30
Chloroform
Concen: 5.175 ug/l
RT: 6.08 min Scan# 463
Delta R.T. 0.01 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Tgt Ion: 83 Resp: 146404
Ion Ratio Lower Upper
83 100
85 69.2 53.8 80.8



Abundance Ion 82.95 (82.65 to 83.65): VDC
Ion 84.95 (84.65 to 85.65): VDC









#31
Cyclohexane
Concen: 5.818 ug/l
RT: 6.35 min Scan# 490
Delta R.T. 0.01 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC005

Tgt Ion: 56 Resp: 157752
Ion Ratio Lower Upper
56 100
69 34.9 22.8 34.2#
84 86.0 64.9 97.3




Abundance Ion 55.95 (55.65 to 56.65): VDC
Ion 69.00 (68.70 to 69.70): VDC
Ion 84.05 (83.75 to 84.75): VDC






#32
1,1,1-Trichloroethane
Concen: 5.092 ug/l
RT: 6.27 min Scan# 482
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Tgt Ion: 97 Resp: 119063
Ion Ratio Lower Upper
97 100
99 62.2 51.4 77.0
61 45.7 35.5 53.3



Abundance Ion 96.85 (96.55 to 97.55): VDC
Ion 98.85 (98.55 to 99.55): VDC
Ion 60.95 (60.65 to 61.65): VDC









#33
 1,2-Dichloroethane-d4
 Concen: 5.092 ug/l
 RT: 6.76 min Scan# 532
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC005


Tgt Ion: 65 Resp: 63158
 Ion Ratio Lower Upper
 65 100
 67 52.2 0.0 106.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 7.42 min Scan# 599
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Tgt Ion: 114 Resp: 2257213
 Ion Ratio Lower Upper
 114 100
 63 17.8 0.0 36.0
 88 18.0 0.0 36.2









#35
 Dibromofluoromethane
 Concen: 50.000 ug/l
 RT: 6.30 min Scan# 485
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC005


Tgt Ion:113 Resp: 86501
 Ion Ratio Lower Upper
 113 100
 111 95.9 81.5 122.3
 192 19.9 17.4 26.0



#36
 1,1-Dichloropropene
 Concen: 5.261 ug/l
 RT: 6.49 min Scan# 505
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Tgt Ion: 75 Resp: 116179
 Ion Ratio Lower Upper
 75 100
 110 36.6 19.0 57.0
 77 31.6 24.9 37.3









#37
Ethyl Acetate
Concen: 4.858 ug/l
RT: 5.68 min Scan# 422
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 74077
Ion Ratio Lower Upper
43 100
61 14.5 10.5 15.7
70 8.1 6.3 9.5



#38
Carbon Tetrachloride
Concen: 4.959 ug/l
RT: 6.47 min Scan# 503
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Tgt Ion: 117 Resp: 101705
Ion Ratio Lower Upper
117 100
119 93.4 77.8 116.6
121 32.5 24.8 37.2









#39
Methylcyclohexane
Concen: 5.062 ug/l
RT: 8.02 min Scan# 660
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC005


Tgt Ion: 83 Resp: 126848
Ion Ratio Lower Upper
83 100
55 88.0 70.5 105.7
98 46.9 39.4 59.2



#40
Benzene
Concen: 5.259 ug/l
RT: 6.77 min Scan# 533
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Tgt Ion: 78 Resp: 296402
Ion Ratio Lower Upper
78 100
77 23.3 19.0 28.4





#41

Methacrylonitrile

Concen: 10.486 ug/l

RT: 5.93 min Scan# 448

Delta R.T. 0.02 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC005

Tgt Ion: 41 Resp: 85076


Ion Ratio Lower Upper

41 100

39 40.4 36.0 54.0

67 3.2 13.9 20.9#

52 2.2 6.1 9.1#






Abundance Ion 40.95 (40.65 to 41.65): VDC

Ion 38.95 (38.65 to 39.65): VDC

Ion 67.00 (66.70 to 67.70): VDC

Ion 52.00 (51.70 to 52.70): VDC



#42

1,2-Dichloroethane

Concen: 4.889 ug/l

RT: 6.87 min Scan# 543

Delta R.T. 0.00 min

Lab File: VD060099.D


Acq: 1 Oct 2018 16:24

Tgt Ion: 62 Resp: 82647

Ion Ratio Lower Upper



62 100


98 10.5 0.0 20.6



Abundance Ion 61.95 (61.65 to 62.65): VDC

Ion 97.85 (97.55 to 98.55): VDC





#43

Isopropyl Acetate

Concen: 4.569 ug/l

RT: 8.32 min Scan# 691

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC005





Tgt Ion: 43 Resp: 92324

Ion Ratio Lower Upper

43 100

61 26.7 21.4 32.2

87 0.0 0.2 0.2#



#44

Trichloroethene

Concen: 5.067 ug/l

RT: 7.71 min Scan# 629

Delta R.T. 0.00 min

Lab File: VD060099.D




Acq: 1 Oct 2018 16:24


Tgt Ion:130 Resp: 88702

Ion Ratio Lower Upper

130 100

95 105.3 0.0 202.2





#45

1,2-Dichloropropane

Concen: 4.928 ug/l

RT: 8.08 min Scan# 666

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :


VSTDIC005

Tgt Ion: 63 Resp: 74351

Ion Ratio Lower Upper




63 100

65 30.9 24.1 36.1



Abundance Ion 62.95 (62.65 to 63.65): VDC

Ion 64.95 (64.65 to 65.65): VDC



#46

Dibromomethane

Concen: 4.911 ug/l

RT: 8.20 min Scan# 678

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24


Tgt Ion: 93 Resp: 48160

Ion Ratio Lower Upper

93 100

95 83.8 63.8 95.6



174 112.0 92.9 139.3




Abundance Ion 92.90 (92.60 to 93.60): VDC

Ion 94.90 (94.60 to 95.60): VDC

Ion 173.80 (173.50 to 174.50): V





#47

Bromodichloromethane

Concen: 4.778 ug/l

RT: 8.47 min Scan# 706

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC005


Tgt Ion: 83 Resp: 101276

Ion Ratio Lower Upper




83 100

85 65.2 52.2 78.2

127 8.7 7.4 11.0



Abundance Ion 82.85 (82.55 to 83.55): VDC
Ion 84.95 (84.65 to 85.65): VDC
Ion 126.80 (126.50 to 127.50): V



#48

Methyl methacrylate

Concen: 4.685 ug/l

RT: 8.23 min Scan# 681

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24


Tgt Ion: 41 Resp: 52777

Ion Ratio Lower Upper



41 100


69 56.4 49.7 74.5

39 55.4 45.2 67.8



Abundance Ion 40.95 (40.65 to 41.65): VDC
Ion 69.00 (68.70 to 69.70): VDC
Ion 38.95 (38.65 to 39.65): VDC









#49
1,4-Dioxane
Concen: 92.168 ug/l
RT: 8.21 min Scan# 679
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC005


Tgt Ion: 88 Resp: 10332
Ion Ratio Lower Upper
88 100
43 48.7 33.7 50.5
58 53.5 49.1 73.7



#50
Toluene-d8
Concen: 92.168 ug/l
RT: 9.44 min Scan# 804
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Tgt Ion: 98 Resp: 241825
Ion Ratio Lower Upper
98 100
100 68.8 55.4 83.0





#51

4-Methyl-2-Pentanone

Concen: 22.096 ug/l

RT: 9.33 min Scan# 793

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :


VSTDIC005

Tgt Ion: 43 Resp: 306288

Ion Ratio Lower Upper




43 100

58 32.9 27.2 40.8



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 57.95 (57.65 to 58.65): VDC



#52

Toluene

Concen: 5.120 ug/l

RT: 9.54 min Scan# 814

Delta R.T. 0.01 min

Lab File: VD060099.D


Acq: 1 Oct 2018 16:24

Tgt Ion: 92 Resp: 186233

Ion Ratio Lower Upper



92 100


91 171.9 131.9 197.9



Abundance Ion 92.00 (91.70 to 92.70): VDC

Ion 91.00 (90.70 to 91.70): VDC









#53
 t-1,3-Dichloropropene
 Concen: 4.576 ug/l
 RT: 9.90 min Scan# 851
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC005


Tgt Ion: 75 Resp: 90587
 Ion Ratio Lower Upper
 75 100
 77 35.5 26.2 39.2



#54
 cis-1,3-Dichloropropene
 Concen: 4.756 ug/l
 RT: 9.08 min Scan# 768
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Tgt Ion: 75 Resp: 112344
 Ion Ratio Lower Upper
 75 100
 77 33.1 25.7 38.5
 39 66.9 48.2 72.4





#55

1,1,2-Trichloroethane

Concen: 4.832 ug/l

RT: 10.16 min Scan# 878

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC005

Tgt Ion: 97 Resp: 60958


Ion Ratio Lower Upper

97 100

83 84.8 67.0 100.6

85 59.0 43.9 65.9

99 56.5 49.1 73.7






Abundance Ion 96.85 (96.55 to 97.55): VDC

Ion 82.95 (82.65 to 83.65): VDC

Ion 84.95 (84.65 to 85.65): VDC

Ion 98.85 (98.55 to 99.55): VDC



#56

Ethyl methacrylate

Concen: 4.448 ug/l

RT: 10.02 min Scan# 863

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24


Tgt Ion: 69 Resp: 61854

Ion Ratio Lower Upper

69 100

41 86.4 66.6 100.0



39 47.2 36.0 54.0




Abundance Ion 69.00 (68.70 to 69.70): VDC

Ion 41.05 (40.75 to 41.75): VDC

Ion 38.95 (38.65 to 39.65): VDC





#57

1,3-Dichloropropane

Concen: 4.905 ug/l

RT: 10.37 min Scan# 899

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :


VSTDIC005

Tgt Ion: 76 Resp: 91142

Ion Ratio Lower Upper




76 100

78 30.5 26.2 39.2



Abundance Ion 75.95 (75.65 to 76.65): VDC

Ion 77.95 (77.65 to 78.65): VDC



#58

2-Chloroethyl Vinyl ether

Concen: 28.294 ug/l

RT: 8.91 min Scan# 750

Delta R.T. 0.00 min

Lab File: VD060099.D


Acq: 1 Oct 2018 16:24

Tgt Ion: 63 Resp: 194032

Ion Ratio Lower Upper



63 100


106 25.9 22.6 34.0



Abundance Ion 62.95 (62.65 to 63.65): VDC

Ion 105.95 (105.65 to 106.65): V





#59

2-Hexanone

Concen: 21.437 ug/l

RT: 10.47 min Scan# 909

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :


VSTDIC005

Tgt Ion: 43 Resp: 214434

Ion Ratio Lower Upper




43 100

58 46.3 23.3 69.9



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 57.95 (57.65 to 58.65): VDC



#60

Dibromochloromethane

Concen: 4.498 ug/l

RT: 10.63 min Scan# 925

Delta R.T. 0.00 min

Lab File: VD060099.D


Acq: 1 Oct 2018 16:24

Tgt Ion: 129 Resp: 68462

Ion Ratio Lower Upper



129 100


127 79.3 37.4 112.2



Abundance Ion 128.80 (128.50 to 129.50): V

Ion 126.80 (126.50 to 127.50): V





#61

1,2-Dibromoethane

Concen: 4.697 ug/l

RT: 10.75 min Scan# 937

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :


VSTDIC005

Tgt Ion: 107 Resp: 61159

Ion Ratio Lower Upper




107 100

109 97.5 77.1 115.7



Abundance Ion 106.90 (106.60 to 107.60): V

Ion 108.90 (108.60 to 109.60): V



#62

4-Bromofluorobenzene

Concen: 4.697 ug/l

RT: 12.42 min Scan# 1107

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24


Tgt Ion: 95 Resp: 91865

Ion Ratio Lower Upper

95 100

174 110.8 0.0 200.2



176 102.9 0.0 195.6




Abundance Ion 95.00 (94.70 to 95.70): V

Ion 173.90 (173.60 to 174.60): V

Ion 175.90 (175.60 to 176.60): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.27 min Scan# 990
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC005

Tgt Ion:117 Resp: 1805814
Ion Ratio Lower Upper
117 100
82 55.2 43.4 65.2
119 32.5 26.0 39.0




Abundance Ion 116.95 (116.65 to 117.65): V
Ion 82.05 (81.75 to 82.75): V
Ion 118.95 (118.65 to 119.65): V






#64
Tetrachloroethene
Concen: 5.175 ug/l
RT: 10.23 min Scan# 885
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Tgt Ion:164 Resp: 81400
Ion Ratio Lower Upper
164 100
166 131.9 105.4 158.0
129 81.5 65.7 98.5
131 80.3 62.9 94.3



Abundance Ion 163.85 (163.55 to 164.55): V
Ion 165.85 (165.55 to 166.55): V
Ion 128.80 (128.50 to 129.50): V
Ion 130.80 (130.50 to 131.50): V









#65
Chlorobenzene
Concen: 5.273 ug/l
RT: 11.30 min Scan# 993
Delta R.T. -0.01 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC005

Tgt Ion:112 Resp: 202494
Ion Ratio Lower Upper
112 100
114 32.5 26.0 39.0




Abundance Ion 111.90 (111.60 to 112.60): V
Ion 114.00 (113.70 to 114.70): V






#66
1,1,1,2-Tetrachloroethane
Concen: 5.109 ug/l
RT: 11.40 min Scan# 1004
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Tgt Ion:131 Resp: 66005
Ion Ratio Lower Upper
131 100
133 98.3 47.4 142.3
119 70.3 36.1 108.3



Abundance Ion 130.90 (130.60 to 131.60): V
Ion 132.90 (132.60 to 133.60): V
Ion 118.85 (118.55 to 119.55): V









#67
Ethyl Benzene
Concen: 5.664 ug/l
RT: 11.41 min Scan# 1005
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC005


Tgt Ion: 91 Resp: 349227
Ion Ratio Lower Upper
91 100
106 29.7 24.4 36.6



#68
m/p-Xylenes
Concen: 11.122 ug/l
RT: 11.55 min Scan# 1019
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Tgt Ion: 106 Resp: 252561
Ion Ratio Lower Upper
106 100
91 216.2 160.0 240.0









#69
o-Xylene
Concen: 5.413 ug/l
RT: 11.93 min Scan# 1057
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC005

Tgt Ion:106 Resp: 122586
Ion Ratio Lower Upper
106 100
91 207.6 102.5 307.5




Abundance Ion 106.05 (105.75 to 106.75): V
Ion 91.00 (90.70 to 91.70): V






#70
Styrene
Concen: 5.323 ug/l
RT: 11.95 min Scan# 1059
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Tgt Ion:104 Resp: 202422
Ion Ratio Lower Upper
104 100
78 42.8 32.2 48.4
103 46.7 37.4 56.0



Abundance Ion 104.05 (103.75 to 104.75): V
Ion 77.95 (77.65 to 78.65): V
Ion 103.05 (102.75 to 103.75): V





#71

Bromoform

Concen: 4.283 ug/l

RT: 12.10 min Scan# 1075

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC005


Tgt Ion:173 Resp: 47362

Ion Ratio Lower Upper




173 100

175 51.2 24.1 72.4

254 9.0 7.4 11.0



Abundance Ion 172.80 (172.50 to 173.50): V
Ion 174.80 (174.50 to 175.50): V
Ion 253.70 (253.40 to 254.40): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.36 min Scan# 1203

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24


Tgt Ion:152 Resp: 952202

Ion Ratio Lower Upper



152 100


115 52.2 27.1 81.2

150 155.3 0.0 311.8



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 115.00 (114.70 to 115.70): V
Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 5.438 ug/l

RT: 12.27 min Scan# 1092

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :


VSTDIC005

Tgt Ion: 105 Resp: 336457

Ion Ratio Lower Upper

105 100

120 24.9 12.6 37.6



Abundance Ion 105.05 (104.75 to 105.75): V

Ion 120.05 (119.75 to 120.75): V

12.27

250000

200000

150000



100000

50000

0

Time-->

12.20 12.30 12.40



#74

N-ethyl acetate

Concen: 4.859 ug/l

RT: 12.12 min Scan# 1077

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Tgt Ion: 43 Resp: 118321


Ion Ratio Lower Upper

43 100

70 25.5 22.1 33.1

55 17.2 14.1 21.1

61 22.4 18.4 27.6



Abundance Ion 42.95 (42.65 to 43.65): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 60.95 (60.65 to 61.65): V

12.12


100000


50000

0

Time-->

12.10 12.20





#75

1,1,2,2-Tetrachloroethane

Concen: 5.096 ug/l

RT: 12.56 min Scan# 1121

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC005


Tgt Ion: 83 Resp: 71475

Ion Ratio Lower Upper

83 100

131 9.3 4.5 13.4




85 65.1 33.3 99.9



Abundance Ion 82.95 (82.65 to 83.65): VDC

Ion 130.90 (130.60 to 131.60): VDC

Ion 84.95 (84.65 to 85.65): VDC



#76

1,2,3-Trichloropropane

Concen: 4.765 ug/l

RT: 12.59 min Scan# 1124

Delta R.T. 0.00 min

Lab File: VD060099.D


Acq: 1 Oct 2018 16:24

Tgt Ion: 75 Resp: 68200

Ion Ratio Lower Upper



75 100


77 37.0 16.3 48.9



Abundance Ion 75.00 (74.70 to 75.70): VDC

Ion 76.95 (76.65 to 77.65): VDC





#77

Bromobenzene

Concen: 5.430 ug/l

RT: 12.53 min Scan# 1118

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC005


Tgt Ion: 156 Resp: 97342

Ion Ratio Lower Upper

156 100

77 133.1 66.9 200.7




158 93.9 49.1 147.4



Abundance Ion 155.90 (155.60 to 156.60): V

Ion 76.95 (76.65 to 77.65): VDC

Ion 157.85 (157.55 to 158.55): V



#78

n-propylbenzene

Concen: 5.620 ug/l

RT: 12.63 min Scan# 1128

Delta R.T. 0.00 min

Lab File: VD060099.D


Acq: 1 Oct 2018 16:24

Tgt Ion: 91 Resp: 435918

Ion Ratio Lower Upper



91 100


120 19.1 10.0 30.0



Abundance Ion 91.00 (90.70 to 91.70): VDC

Ion 120.05 (119.75 to 120.75): V









#79
 2-Chlorotoluene
 Concen: 5.305 ug/l
 RT: 12.69 min Scan# 1135
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC005


Tgt Ion: 91 Resp: 231837
 Ion Ratio Lower Upper
 91 100
 126 32.5 16.8 50.3



#80
 1,3,5-Trimethylbenzene
 Concen: 5.677 ug/l
 RT: 12.78 min Scan# 1144
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Tgt Ion: 105 Resp: 269106
 Ion Ratio Lower Upper
 105 100
 120 47.7 23.7 71.1





#81

trans-1,4-Dichloro-2-butene

Concen: 4.151 ug/l

RT: 12.33 min Scan# 1098

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC005


Tgt Ion: 75 Resp: 17901

Ion Ratio Lower Upper

75 100

53 96.8 73.0 109.6




89 36.5 32.0 48.0



Abundance Ion 75.00 (74.70 to 75.70): VDC

Ion 53.00 (52.70 to 53.70): VDC

Ion 89.00 (88.70 to 89.70): VDC



#82

4-Chlorotoluene

Concen: 5.827 ug/l

RT: 12.80 min Scan# 1146

Delta R.T. 0.00 min

Lab File: VD060099.D


Acq: 1 Oct 2018 16:24

Tgt Ion: 91 Resp: 279993

Ion Ratio Lower Upper



91 100


126 34.3 17.1 51.2



Abundance Ion 91.00 (90.70 to 91.70): VDC

Ion 125.95 (125.65 to 126.65): V









#83
 tert-Butylbenzene
 Concen: 5.293 ug/l
 RT: 13.04 min Scan# 1170
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC005


Tgt Ion:119 Resp: 291291
 Ion Ratio Lower Upper
 119 100
 91 69.8 34.8 104.4
 134 24.7 11.8 35.4



#84
 1,2,4-Trimethylbenzene
 Concen: 5.423 ug/l
 RT: 13.09 min Scan# 1175
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Tgt Ion:105 Resp: 275179
 Ion Ratio Lower Upper
 105 100
 120 45.6 23.4 70.0









#85
 sec-Butylbenzene
 Concen: 5.508 ug/l
 RT: 13.22 min Scan# 1188
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC005


Tgt Ion:105 Resp: 357723
 Ion Ratio Lower Upper
 105 100
 134 18.4 8.8 26.2



#86
 p-Isopropyltoluene
 Concen: 5.552 ug/l
 RT: 13.33 min Scan# 1200
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Tgt Ion:119 Resp: 291827
 Ion Ratio Lower Upper
 119 100
 134 24.8 12.3 36.8
 91 22.9 11.8 35.4









#87
 1,3-Dichlorobenzene
 Concen: 5.456 ug/l
 RT: 13.30 min Scan# 1197
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC005

Tgt Ion:146 Resp: 169772
 Ion Ratio Lower Upper
 146 100
 111 34.5 18.1 54.3
 148 64.4 32.3 96.9




Abundance Ion 145.95 (145.65 to 146.65): V
 Ion 110.90 (110.60 to 111.60): V
 Ion 147.90 (147.60 to 148.60): V






#88
 1,4-Dichlorobenzene
 Concen: 5.465 ug/l
 RT: 13.38 min Scan# 1205
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Tgt Ion:146 Resp: 172340
 Ion Ratio Lower Upper
 146 100
 111 32.7 17.8 53.3
 148 63.6 32.4 97.2



Abundance Ion 145.95 (145.65 to 146.65): V
 Ion 110.90 (110.60 to 111.60): V
 Ion 147.90 (147.60 to 148.60): V





#89

n-Butylbenzene

Concen: 5.900 ug/l

RT: 13.64 min Scan# 1231

Delta R.T. 0.00 min

Lab File: VD060099.D

Acq: 1 Oct 2018 16:24

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC005





Tgt Ion: 91 Resp: 302060

Ion Ratio Lower Upper

91 100

92 60.2 30.1 90.3

134 22.4 11.6 34.7



#90

Hexachloroethane

Concen: 4.438 ug/l

RT: 13.85 min Scan# 1252

Delta R.T. 0.00 min

Lab File: VD060099.D




Acq: 1 Oct 2018 16:24


Tgt Ion: 117 Resp: 57682

Ion Ratio Lower Upper

117 100

201 65.5 34.8 104.5









#91
 1,2-Dichlorobenzene
 Concen: 5.929 ug/l
 RT: 13.65 min Scan# 1232
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC005


Tgt Ion: 146 Resp: 152262
 Ion Ratio Lower Upper
 146 100
 111 41.3 21.4 64.3
 148 62.3 32.2 96.6



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 4.133 ug/l
 RT: 14.22 min Scan# 1290
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Tgt Ion: 75 Resp: 8855
 Ion Ratio Lower Upper
 75 100
 155 103.5 55.9 167.7
 157 124.4 73.3 219.8









#93
 1,2,4-Trichlorobenzene
 Concen: 5.215 ug/l
 RT: 14.78 min Scan# 1347
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC005

Tgt Ion:180 Resp: 115645
 Ion Ratio Lower Upper
 180 100
 182 93.2 47.3 141.8
 145 24.0 13.0 38.9




Abundance Ion 179.85 (179.55 to 180.55): V
 Ion 181.85 (181.55 to 182.55): V
 Ion 144.95 (144.65 to 145.65): V






#94
 Hexachlorobutadiene
 Concen: 5.108 ug/l
 RT: 14.87 min Scan# 1356
 Delta R.T. 0.00 min
 Lab File: VD060099.D
 Acq: 1 Oct 2018 16:24

Tgt Ion:225 Resp: 75044
 Ion Ratio Lower Upper
 225 100
 223 60.5 31.8 95.4
 227 67.6 32.0 96.0



Abundance Ion 224.75 (224.45 to 225.45): V
 Ion 222.75 (222.45 to 223.45): V
 Ion 226.75 (226.45 to 227.45): V









#95
Naphthalene
Concen: 4.821 ug/l
RT: 14.96 min Scan# 1365
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Instrument :
MSVOA_D
ClientSampleId :
VSTDIC005

Tgt Ion:128 Resp: 164874
Ion Ratio Lower Upper
128 100
127 12.2 9.9 14.9
129 10.9 8.8 13.2




Abundance Ion 128.00 (127.70 to 128.70): V
Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V



#96
1,2,3-Trichlorobenzene
Concen: 5.169 ug/l
RT: 15.11 min Scan# 1380
Delta R.T. 0.00 min
Lab File: VD060099.D
Acq: 1 Oct 2018 16:24

Tgt Ion:180 Resp: 100398
Ion Ratio Lower Upper
180 100
182 96.0 48.4 145.0
145 29.6 14.5 43.6



Abundance Ion 179.85 (179.55 to 180.55): V
Ion 181.85 (181.55 to 182.55): V
Ion 144.95 (144.65 to 145.65): V

