

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD100120\
 Data File : VD067020.D
 Acq On : 02 Oct 2020 05:10
 Operator : VA/SY
 Sample : L3872-13
 Misc : 7.52G/5.00ml/MSVOA D/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BU-03-093020

Quant Time: Oct 02 06:08:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 15:19:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.97	168	15552	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	23354	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	24527	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	10894	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	12567	83.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	167.78%	
35) Dibromofluoromethane	7.92	113	6999	45.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.56%	
50) Toluene-d8	10.34	98	27191	48.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.62%	
62) 4-Bromofluorobenzene	12.63	95	9805	51.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.68%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD100120\
Data File : VD067020.D
Acq On : 02 Oct 2020 05:10
Operator : VA/SY
Sample : L3872-13
Misc : 7.52G/5.00ml/MSVOA D/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BU-03-093020

Quant Time: Oct 02 06:08:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 15:19:17 2020
Response via : Initial Calibration

