

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD100120\
 Data File : VD067024.D
 Acq On : 02 Oct 2020 07:00
 Operator : VA/SY
 Sample : L4213-07
 Misc : 5.06G/5.00ml/MSVOA D/SOIL
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B-16(9-11)

Manual Integrations
 APPROVED

MMDadoda
 10/2/2020 3:13:44 PM

Quant Time: Oct 02 07:27:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 15:19:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	380252	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	633541	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	613842	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	312003	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	212757	58.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.18%	
35) Dibromofluoromethane	7.92	113	213948	51.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.16%	
50) Toluene-d8	10.34	98	765198	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
62) 4-Bromofluorobenzene	12.63	95	288125	56.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.32%	
Target Compounds						
16) Acetone	4.17	43	4323m	7.732	ug/l	Qvalue
86) p-Isopropyltoluene	13.52	119	146261	7.432	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD100120\
Data File : VD067024.D
Acq On : 02 Oct 2020 07:00
Operator : VA/SY
Sample : L4213-07
Misc : 5.06G/5.00ml/MSVOA D/SOIL
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B-16(9-11)

Manual Integrations
APPROVED

MMDadoda
10/2/2020 3:13:44 PM

Quant Time: Oct 02 07:27:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 15:19:17 2020
Response via : Initial Calibration

