

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD100120\
 Data File : VD067007.D
 Acq On : 01 Oct 2020 22:46
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 02 01:30:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 15:19:17 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.454	0.410	9.7	85	0.00
3 P	Chloromethane	0.398	0.380	4.5	91	0.00
4 C	Vinyl Chloride	0.399	0.384	3.8#	88	0.00
5 T	Bromomethane	0.291	0.302	-3.8	98	0.00
6 T	Chloroethane	0.243	0.239	1.6	89	0.00
7 T	Trichlorofluoromethane	0.887	0.834	6.0	87	0.00
8 T	Diethyl Ether	0.220	0.237	-7.7	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.501	0.483	3.6	89	0.00
10 T	Methyl Iodide	0.487	0.535	-9.9	95	0.00
11 T	Tert butyl alcohol	0.041	0.029	29.3#	92	0.00
12 CM	1,1-Dichloroethene	0.471	0.458	2.8#	88	0.00
13 T	Acrolein	0.032	0.029	9.4	85	0.00
14 T	Allyl chloride	0.612	0.641	-4.7	91	0.00
15 T	Acrylonitrile	0.081	0.085	-4.9	92	0.00
16 T	Acetone	0.074	0.069	6.8	81	0.00
17 T	Carbon Disulfide	1.447	1.427	1.4	89	0.00
18 T	Methyl Acetate	0.185	0.212	-14.6	101	0.00
19 T	Methyl tert-butyl Ether	0.860	0.928	-7.9	90	0.00
20 T	Methylene Chloride	0.670	0.639	4.6	103	0.00
21 T	trans-1,2-Dichloroethene	0.536	0.551	-2.8	91	0.00
22 T	Diisopropyl ether	1.118	1.225	-9.6	93	0.00
23 T	Vinyl Acetate	0.633	0.696	-10.0	89	0.00
24 P	1,1-Dichloroethane	0.863	0.886	-2.7	91	0.00
25 T	2-Butanone	0.104	0.107	-2.9	92	0.00
26 T	2,2-Dichloropropane	0.828	0.725	12.4	82	0.00
27 T	cis-1,2-Dichloroethene	0.573	0.607	-5.9	93	0.00
28 T	Bromochloromethane	0.333	0.368	-10.5	92	0.00
29 T	Tetrahydrofuran	0.060	0.063	-5.0	90	0.00
30 C	Chloroform	0.936	0.975	-4.2#	94	0.00
31 T	Cyclohexane	0.723	0.661	8.6	85	0.00
32 T	1,1,1-Trichloroethane	0.867	0.881	-1.6	92	0.00
33 S	1,2-Dichloroethane-d4	0.482	0.510	-5.8	91	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.327	0.341	-4.3	94	0.00
36 T	1,1-Dichloropropene	0.466	0.446	4.3	90	0.00
37 T	Ethyl Acetate	0.161	0.155	3.7	90	0.00
38 T	Carbon Tetrachloride	0.504	0.480	4.8	89	0.00
39 T	Methylcyclohexane	0.493	0.472	4.3	84	0.00
40 TM	Benzene	1.294	1.287	0.5	93	0.00
41 T	Methacrylonitrile	0.087	0.090	-3.4	89	0.00
42 TM	1,2-Dichloroethane	0.362	0.353	2.5	92	0.00
43 T	Isopropyl Acetate	0.297	0.292	1.7	89	0.00
44 TM	Trichloroethene	0.387	0.378	2.3	93	0.00
45 C	1,2-Dichloropropane	0.304	0.316	-3.9#	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD100120\
 Data File : VD067007.D
 Acq On : 01 Oct 2020 22:46
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 02 01:30:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 15:19:17 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.175	0.179	-2.3	92	0.00
47 T	Bromodichloromethane	0.454	0.460	-1.3	94	0.00
48 T	Methyl methacrylate	0.152	0.144	5.3	81	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	85	0.00
50 S	Toluene-d8	1.193	1.247	-4.5	91	0.00
51 T	4-Methyl-2-Pentanone	0.148	0.151	-2.0	89	0.00
52 CM	Toluene	0.834	0.848	-1.7#	92	0.00
53 T	t-1,3-Dichloropropene	0.406	0.409	-0.7	90	0.00
54 T	cis-1,3-Dichloropropene	0.496	0.499	-0.6	92	0.00
55 T	1,1,2-Trichloroethane	0.243	0.237	2.5	90	0.00
56 T	Ethyl methacrylate	0.251	0.272	-8.4	90	0.00
57 T	1,3-Dichloropropane	0.394	0.411	-4.3	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.140	0.154	-10.0	95	0.00
59 T	2-Hexanone	0.099	0.102	-3.0	88	0.00
60 T	Dibromochloromethane	0.324	0.326	-0.6	91	0.00
61 T	1,2-Dibromoethane	0.233	0.238	-2.1	91	0.00
62 S	4-Bromofluorobenzene	0.405	0.415	-2.5	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.364	0.347	4.7	92	0.00
65 PM	Chlorobenzene	0.994	0.975	1.9	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.378	0.380	-0.5	95	0.00
67 C	Ethyl Benzene	1.704	1.721	-1.0#	92	0.00
68 T	m/p-Xylenes	0.664	0.669	-0.8	92	0.00
69 T	o-Xylene	0.594	0.617	-3.9	92	0.00
70 T	Styrene	1.020	1.080	-5.9	94	0.00
71 P	Bromoform	0.196	0.196	0.0	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	3.395	3.529	-3.9	90	0.00
74 T	N-amyl acetate	0.610	0.621	-1.8	90	0.00
75 P	1,1,2,2-Tetrachloroethane	0.572	0.575	-0.5	92	0.00
76 T	1,2,3-Trichloropropane	0.436	0.389	10.8	79	0.00
77 T	Bromobenzene	0.820	0.862	-5.1	97	0.00
78 T	n-propylbenzene	4.012	4.072	-1.5	88	0.00
79 T	2-Chlorotoluene	2.296	2.333	-1.6	91	0.00
80 T	1,3,5-Trimethylbenzene	2.868	2.998	-4.5	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.175	0.172	1.7	89	0.00
82 T	4-Chlorotoluene	2.442	2.433	0.4	90	0.00
83 T	tert-Butylbenzene	2.415	2.576	-6.7	94	0.00
84 T	1,2,4-Trimethylbenzene	2.863	2.990	-4.4	92	0.00
85 T	sec-Butylbenzene	3.402	3.460	-1.7	89	0.00
86 T	p-Isopropyltoluene	3.154	3.175	-0.7	88	0.00
87 T	1,3-Dichlorobenzene	1.597	1.572	1.6	91	0.00
88 T	1,4-Dichlorobenzene	1.597	1.533	4.0	90	0.00
89 T	n-Butylbenzene	2.882	2.789	3.2	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD100120\
Data File : VD067007.D
Acq On : 01 Oct 2020 22:46
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Oct 02 01:30:01 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 15:19:17 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.552	0.557	-0.9	91	0.00
91 T	1,2-Dichlorobenzene	1.395	1.382	0.9	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.092	0.091	1.1	89	0.00
93 T	1,2,4-Trichlorobenzene	0.902	0.909	-0.8	88	0.00
94 T	Hexachlorobutadiene	0.554	0.530	4.3	87	0.00
95 T	Naphthalene	1.505	1.592	-5.8	88	0.00
96 T	1,2,3-Trichlorobenzene	0.783	0.792	-1.1	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6