

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD100318\
 Data File : VD060115.D
 Acq On : 3 Oct 2018 13:29
 Operator : VA/AP
 Sample : VD1003SBL01
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1003SBL01

Quant Time: Oct 04 04:43:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 06:56:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.39	168	1645294	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	7.44	114	2238857	50.00	ug/l	0.02
63) Chlorobenzene-d5	11.28	117	1822112	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	13.37	152	1004344	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.78	65	680779	54.59	ug/l	0.02
Spiked Amount 50.000			Recovery	=	109.18%	
35) Dibromofluoromethane	6.32	113	862498	50.73	ug/l	0.02
Spiked Amount 50.000			Recovery	=	101.46%	
50) Toluene-d8	9.45	98	2452242	50.58	ug/l	0.02
Spiked Amount 50.000			Recovery	=	101.16%	
62) 4-Bromofluorobenzene	12.43	95	944296	52.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.04%	
Target Compounds						
20) Methylene Chloride	3.71	84	145795	4.250	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD100318\
Data File : VD060115.D
Acq On : 3 Oct 2018 13:29
Operator : VA/AP
Sample : VD1003SBL01
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1003SBL01

Quant Time: Oct 04 04:43:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D100118S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 06:56:35 2018
Response via : Initial Calibration

