

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD100318\
 Data File : VD060122.D
 Acq On : 3 Oct 2018 17:09
 Operator : VA/AP
 Sample : J5237-01
 Misc : 4.50µl/5ml/MSVOA D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BSS-UST-1A

Quant Time: Oct 04 04:58:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 06:56:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.40	168	1340576	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	7.43	114	1832780	50.00	ug/l	0.01
63) Chlorobenzene-d5	11.28	117	1505517	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.38	152	808285	50.00	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.77	65	535296	52.68	ug/l	0.01
Spiked Amount 50.000			Recovery	=	105.36%	
35) Dibromofluoromethane	6.31	113	718863	51.65	ug/l	0.01
Spiked Amount 50.000			Recovery	=	103.30%	
50) Toluene-d8	9.45	98	1881133	47.40	ug/l	0.01
Spiked Amount 50.000			Recovery	=	94.80%	
62) 4-Bromofluorobenzene	12.42	95	754521	50.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.56%	
Target Compounds						
20) Methylene Chloride	3.70	84	196049	10.074	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.91	63	12083	2.043	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD100318\
Data File : VD060122.D
Acq On : 3 Oct 2018 17:09
Operator : VA/AP
Sample : J5237-01
Misc : 4.50µl/5ml/MSVOA D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BSS-UST-1A

Quant Time: Oct 04 04:58:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D100118S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 06:56:35 2018
Response via : Initial Calibration

