

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD100319\
 Data File : VD063877.D
 Acq On : 03 Oct 2019 15:59
 Operator : VA/SY
 Sample : VD1003SBL06
 Misc : 5.00g/5.00mL/MSVOA D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1003SBL06

Quant Time: Oct 04 01:49:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100219S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 03 04:49:16 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	191138	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	306594	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	291226	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	145123	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.34	65	85495	45.94	ug/l	0.00
Spiked Amount			Recovery	=	91.88%	
35) Dibromofluoromethane	7.92	113	91432	48.80	ug/l	0.00
Spiked Amount			Recovery	=	97.60%	
50) Toluene-d8	10.34	98	360806	47.44	ug/l	0.00
Spiked Amount			Recovery	=	94.88%	
62) 4-Bromofluorobenzene	12.63	95	112433	42.54	ug/l	0.00
Spiked Amount			Recovery	=	85.08%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD100319\
Data File : VD063877.D
Acq On : 03 Oct 2019 15:59
Operator : VA/SY
Sample : VD1003SBL06
Misc : 5.00µ/5.00mL/MSVOA D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1003SBL06

Quant Time: Oct 04 01:49:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D100219S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 03 04:49:16 2019
Response via : Initial Calibration

