

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100421\
 Data File : VD070551.D
 Acq On : 04 Oct 2021 15:16
 Operator : VA/SY
 Sample : M4042-03
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EO-02-100121

Quant Time: Oct 05 05:00:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	6571	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	11084	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	9687	50.00	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	2948	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	5052	59.88	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.76%
35) Dibromofluoromethane	7.909	113	3844	48.72	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.44%
50) Toluene-d8	10.332	98	12435	43.41	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	86.82%
62) 4-Bromofluorobenzene	12.620	95	3570	35.66	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	71.32%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100421\
Data File : VD070551.D
Acq On : 04 Oct 2021 15:16
Operator : VA/SY
Sample : M4042-03
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EO-02-100121

Quant Time: Oct 05 05:00:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

