

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100421\
 Data File : VD070558.D
 Acq On : 04 Oct 2021 18:34
 Operator : VA/SY
 Sample : M4046-02
 Misc : 5.10G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-1-(1.5-2.0)

Quant Time: Oct 05 05:01:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	368182	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	779742	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	749557	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	310048	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	284162	60.11	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.22%
35) Dibromofluoromethane	7.914	113	275838	49.70	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.40%
50) Toluene-d8	10.338	98	1010791	50.16	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.32%
62) 4-Bromofluorobenzene	12.626	95	321185	45.60	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.20%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100421\
Data File : VD070558.D
Acq On : 04 Oct 2021 18:34
Operator : VA/SY
Sample : M4046-02
Misc : 5.10G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-1-(1.5-2.0)

Quant Time: Oct 05 05:01:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

