

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100421\
 Data File : VD070560.D
 Acq On : 04 Oct 2021 19:31
 Operator : VA/SY
 Sample : M4046-04
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-1-(4.5-5.0)

Quant Time: Oct 05 05:01:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	370265	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	777270	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	733097	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	305372	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	264143	55.56	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.12%
35) Dibromofluoromethane	7.908	113	272522	49.26	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.52%
50) Toluene-d8	10.338	98	1017584	50.66	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.32%
62) 4-Bromofluorobenzene	12.626	95	314834	44.85	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.70%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100421\
Data File : VD070560.D
Acq On : 04 Oct 2021 19:31
Operator : VA/SY
Sample : M4046-04
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-1-(4.5-5.0)

Quant Time: Oct 05 05:01:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

