

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100421\
 Data File : VD070545.D
 Acq On : 04 Oct 2021 12:17
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 05 04:58:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.574	0.475	17.2	85	0.00
3 P	Chloromethane	0.771	0.608	21.1	81	0.00
4 C	Vinyl Chloride	0.787	0.743	5.6#	92	0.00
5 T	Bromomethane	0.580	0.453	21.9	85	0.00
6 T	Chloroethane	0.541	0.485	10.4	93	0.00
7 T	Trichlorofluoromethane	1.041	1.049	-0.8	92	0.00
8 T	Diethyl Ether	0.288	0.284	1.4	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.646	0.645	0.2	90	0.00
10 T	Methyl Iodide	0.564	0.564	0.0	75	0.00
11 T	Tert butyl alcohol	0.073	0.031	57.5#	52	0.00
12 CM	1,1-Dichloroethene	0.582	0.574	1.4#	87	0.00
13 T	Acrolein	0.032	0.015	53.1#	43#	0.00
14 T	Allyl chloride	0.855	0.828	3.2	87	0.00
15 T	Acrylonitrile	0.137	0.131	4.4	86	0.00
16 T	Acetone	0.115	0.135	-17.4	111	0.00
17 T	Carbon Disulfide	2.148	1.781	17.1	76	0.00
18 T	Methyl Acetate	0.344	0.305	11.3	88	0.00
19 T	Methyl tert-butyl Ether	1.258	1.297	-3.1	85	0.00
20 T	Methylene Chloride	0.913	0.715	21.7	83	0.00
21 T	trans-1,2-Dichloroethene	0.693	0.674	2.7	85	0.00
22 T	Diisopropyl ether	1.831	1.811	1.1	86	0.00
23 T	Vinyl Acetate	0.848	0.861	-1.5	88	0.00
24 P	1,1-Dichloroethane	1.295	1.196	7.6	85	0.00
25 T	2-Butanone	0.153	0.154	-0.7	93	0.00
26 T	2,2-Dichloropropane	1.047	1.080	-3.2	96	0.00
27 T	cis-1,2-Dichloroethene	0.744	0.735	1.2	86	0.00
28 T	Bromochloromethane	0.514	0.468	8.9	84	0.00
29 T	Tetrahydrofuran	0.098	0.098	0.0	90	0.00
30 C	Chloroform	1.328	1.228	7.5#	85	0.00
31 T	Cyclohexane	1.106	1.077	2.6	89	0.00
32 T	1,1,1-Trichloroethane	1.118	1.130	-1.1	92	0.00
33 S	1,2-Dichloroethane-d4	0.642	0.544	15.3	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.356	0.332	6.7	80	0.00
36 T	1,1-Dichloropropene	0.534	0.571	-6.9	89	0.00
37 T	Ethyl Acetate	0.200	0.198	1.0	88	0.00
38 T	Carbon Tetrachloride	0.529	0.587	-11.0	93	0.00
39 T	Methylcyclohexane	0.597	0.668	-11.9	87	0.00
40 TM	Benzene	1.566	1.611	-2.9	85	0.00
41 T	Methacrylonitrile	0.105	0.107	-1.9	92	0.00
42 TM	1,2-Dichloroethane	0.419	0.408	2.6	85	0.00
43 T	Isopropyl Acetate	0.370	0.367	0.8	88	0.00
44 TM	Trichloroethene	0.374	0.393	-5.1	85	0.00
45 C	1,2-Dichloropropane	0.411	0.404	1.7#	84	0.00
46 T	Dibromomethane	0.212	0.206	2.8	84	0.00
47 T	Bromodichloromethane	0.540	0.536	0.7	85	0.00
48 T	Methyl methacrylate	0.171	0.182	-6.4	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100421\
 Data File : VD070545.D
 Acq On : 04 Oct 2021 12:17
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 05 04:58:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	89	0.00
50 S	Toluene-d8	1.292	1.299	-0.5	81	0.00
51 T	4-Methyl-2-Pentanone	0.191	0.202	-5.8	89	0.00
52 CM	Toluene	0.946	1.039	-9.8#	86	0.00
53 T	t-1,3-Dichloropropene	0.464	0.493	-6.2	88	0.00
54 T	cis-1,3-Dichloropropene	0.565	0.585	-3.5	88	0.00
55 T	1,1,2-Trichloroethane	0.291	0.292	-0.3	83	0.00
56 T	Ethyl methacrylate	0.318	0.354	-11.3	86	0.00
57 T	1,3-Dichloropropane	0.496	0.506	-2.0	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.153	0.175	-14.4	94	0.00
59 T	2-Hexanone	0.128	0.146	-14.1	94	0.00
60 T	Dibromochloromethane	0.339	0.341	-0.6	82	0.00
61 T	1,2-Dibromoethane	0.268	0.268	0.0	83	0.00
62 S	4-Bromofluorobenzene	0.452	0.449	0.7	82	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
64 T	Tetrachloroethene	0.304	0.337	-10.9	85	0.00
65 PM	Chlorobenzene	1.036	1.107	-6.9	84	0.00
66 T	1,1,1,2-Tetrachloroethane	0.379	0.405	-6.9	85	0.00
67 C	Ethyl Benzene	1.786	2.073	-16.1#	87	0.00
68 T	m/p-Xylenes	0.690	0.816	-18.3	86	0.00
69 T	o-Xylene	0.618	0.736	-19.1	86	0.00
70 T	Styrene	1.080	1.284	-18.9	84	0.00
71 P	Bromoform	0.187	0.200	-7.0	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
73 T	Isopropylbenzene	3.506	4.212	-20.1	90	0.00
74 T	N-amyl acetate	0.764	0.806	-5.5	88	0.00
75 P	1,1,2,2-Tetrachloroethane	0.751	0.745	0.8	85	0.00
76 T	1,2,3-Trichloropropane	0.506	0.474	6.3	78	0.00
77 T	Bromobenzene	0.809	0.860	-6.3	84	0.00
78 T	n-propylbenzene	4.551	5.305	-16.6	89	0.00
79 T	2-Chlorotoluene	2.653	3.003	-13.2	88	0.00
80 T	1,3,5-Trimethylbenzene	3.026	3.595	-18.8	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.202	0.237	-17.3	96	0.00
82 T	4-Chlorotoluene	2.799	3.128	-11.8	86	0.00
83 T	tert-Butylbenzene	2.427	3.006	-23.9	92	0.00
84 T	1,2,4-Trimethylbenzene	2.980	3.571	-19.8	88	0.00
85 T	sec-Butylbenzene	3.858	4.647	-20.5	91	0.00
86 T	p-Isopropyltoluene	3.100	3.763	-21.4	89	0.00
87 T	1,3-Dichlorobenzene	1.684	1.816	-7.8	85	0.00
88 T	1,4-Dichlorobenzene	1.691	1.765	-4.4	84	0.00
89 T	n-Butylbenzene	3.091	3.700	-19.7	91	0.00
90 T	Hexachloroethane	0.710	0.736	-3.7	88	0.00
91 T	1,2-Dichlorobenzene	1.486	1.532	-3.1	82	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.112	0.113	-0.9	88	0.00
93 T	1,2,4-Trichlorobenzene	0.802	0.875	-9.1	86	0.00
94 T	Hexachlorobutadiene	0.492	0.555	-12.8	97	0.00
95 T	Naphthalene	1.458	1.658	-13.7	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100421\
Data File : VD070545.D
Acq On : 04 Oct 2021 12:17
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Oct 05 04:58:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.708	0.755	-6.6	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6