

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100521\
 Data File : VD070579.D
 Acq On : 05 Oct 2021 18:38
 Operator : VA/SY
 Sample : M4043-01
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HR-03-100121

Quant Time: Oct 06 09:23:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	30619	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	56112	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	59841	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	24830	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	33004	91.59	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	183.18%#
35) Dibromofluoromethane	7.920	113	24190	64.84	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	129.68%
50) Toluene-d8	10.332	98	72758	51.02	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.04%
62) 4-Bromofluorobenzene	12.626	95	23359	49.01	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.02%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100521\
Data File : VD070579.D
Acq On : 05 Oct 2021 18:38
Operator : VA/SY
Sample : M4043-01
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HR-03-100121

Quant Time: Oct 06 09:23:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

