

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100521\
 Data File : VD070580.D
 Acq On : 05 Oct 2021 19:06
 Operator : VA/SY
 Sample : M4043-03
 Misc : 5.10G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HR-04-100121

Quant Time: Oct 06 09:24:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.979	168	357267	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	738254	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	691700	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	276248	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	235722	56.06	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.12%
35) Dibromofluoromethane	7.914	113	251840	51.31	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.62%
50) Toluene-d8	10.338	98	942865	50.25	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.50%
62) 4-Bromofluorobenzene	12.626	95	276153	44.03	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	88.06%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100521\
 Data File : VD070580.D
 Acq On : 05 Oct 2021 19:06
 Operator : VA/SY
 Sample : M4043-03
 Misc : 5.10G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HR-04-100121

Quant Time: Oct 06 09:24:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

