

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100521\
 Data File : VD070581.D
 Acq On : 05 Oct 2021 19:34
 Operator : VA/SY
 Sample : M4041-01RE
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 N-1RE

Quant Time: Oct 06 09:24:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	120765	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	228607	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	231104	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	90710	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	111053	78.14	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 156.28%	
35) Dibromofluoromethane	7.914	113	88698	58.36	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 116.72%	
50) Toluene-d8	10.337	98	295817	50.91	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 101.82%	
62) 4-Bromofluorobenzene	12.625	95	91762	47.25	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 94.50%	
Target Compounds						
16) Acetone	4.155	43	49379	204.60	ug/l	96
25) 2-Butanone	7.208	43	10356	32.77	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100521\
Data File : VD070581.D
Acq On : 05 Oct 2021 19:34
Operator : VA/SY
Sample : M4041-01RE
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
N-1RE

Quant Time: Oct 06 09:24:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

