

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD100620\
 Data File : VD067103.D
 Acq On : 06 Oct 2020 13:06
 Operator : VA/SY
 Sample : L4272-12
 Misc : 7.27G/5.00ml/MSVOA D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 GIS-1-(6)

Quant Time: Oct 07 01:35:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 15:19:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	405877	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	696173	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	646621	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	287927	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.33	65	204869	52.40	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	104.80%	
35) Dibromofluoromethane	7.91	113	226747	49.75	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.50%	
50) Toluene-d8	10.34	98	813826	49.00	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	98.00%	
62) 4-Bromofluorobenzene	12.64	95	262879	46.63	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	93.26%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD100620\
Data File : VD067103.D
Acq On : 06 Oct 2020 13:06
Operator : VA/SY
Sample : L4272-12
Misc : 7.27G/5.00ml/MSVOA D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
GIS-1-(6)

Quant Time: Oct 07 01:35:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 15:19:17 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.98 min Scan# 1030

Delta R.T. 0.00 min

Lab File: VD067103.D

Acq: 06 Oct 2020 13:06

Instrument :

MSVOA_D

ClientSampleId :


GIS-1-(6)

Tot Ion:168 Resp: 405877

Ion Ratio Lower Upper

168 100

99 55.2 44.1 66.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC

7.98

200000



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 52.404 ug/l

RT: 8.33 min Scan# 1090

Delta R.T. 0.00 min

Lab File: VD067103.D


Acq: 06 Oct 2020 13:06

Tgt Ion: 65 Resp: 204869

Ion Ratio Lower Upper

65 100

67 51.1 0.0 106.8



Abundance Ion 64.90 (64.60 to 65.60): VDC

Ion 66.90 (66.60 to 67.60): VDC

8.33

100000

80000


60000


40000

20000

0

Time-->





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.87 min Scan# 1181

Delta R.T. 0.01 min

Lab File: VD067103.D

Acq: 06 Oct 2020 13:06

Instrument :

MSVOA_D

ClientSampleId :

GIS-1-(6)


Tot Ion:114 Resp: 696173

Ion Ratio Lower Upper

114 100

63 18.4 0.0 38.4




88 14.9 0.0 31.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): V

Ion 87.90 (87.60 to 88.60): V



#35

Dibromofluoromethane

Concen: 49.750 ug/l

RT: 7.91 min Scan# 1019

Delta R.T. 0.00 min

Lab File: VD067103.D

Acq: 06 Oct 2020 13:06


Tgt Ion:113 Resp: 226747

Ion Ratio Lower Upper

113 100

111 103.2 83.0 124.4



192 21.0 15.5 23.3




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 49.003 ug/l
RT: 10.34 min Scan# 1431
Delta R.T. 0.00 min
Lab File: VD067103.D
Acq: 06 Oct 2020 13:06




Instrument :
MSVOA_D
ClientSampleId :
GIS-1-(6)


Tot Ion: 98 Resp: 813826
Ion Ratio Lower Upper
98 100
100 66.7 52.8 79.2



#62
4-Bromofluorobenzene
Concen: 46.627 ug/l
RT: 12.64 min Scan# 1822
Delta R.T. 0.01 min
Lab File: VD067103.D
Acq: 06 Oct 2020 13:06

Tgt Ion: 95 Resp: 262879
Ion Ratio Lower Upper
95 100
174 84.4 0.0 167.6
176 82.2 0.0 162.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.64 min Scan# 1653
Delta R.T. 0.00 min
Lab File: VD067103.D
Acq: 06 Oct 2020 13:06

Instrument :
MSVOA_D
ClientSampleId :
GIS-1-(6)

Tot Ion	Ratio	Lower	Upper
117	100		
82	53.4	42.6	63.8
119	31.2	26.3	39.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.58 min Scan# 1982
Delta R.T. 0.00 min
Lab File: VD067103.D
Acq: 06 Oct 2020 13:06

Tgt Ion	Ratio	Lower	Upper
152	100		
115	57.2	29.6	88.8
150	155.6	0.0	344.8

