


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD100620\
 Data File : VD067095.D
 Acq On : 06 Oct 2020 09:15
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
 Title : SW846 8260
 Last Update : Thu Oct 01 15:19:17 2020



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	42882	PASS
75	95	30	60	48.4	124488	PASS
95	95	100	100	100.0	257216	PASS
96	95	5	9	6.5	16651	PASS
173	174	0.00	2	0.4	734	PASS
174	95	50	100	81.0	208469	PASS
175	174	5	9	7.8	16188	PASS
176	174	95	101	96.6	201429	PASS
177	176	5	9	6.5	13009	PASS