

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100621\
 Data File : VD070585.D
 Acq On : 06 Oct 2021 11:19
 Operator : VA/SY
 Sample : VD1006SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1006SBL01

Quant Time: Oct 07 06:58:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	386024	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	776026	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	725873	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	295223	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	254040	55.919	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.840%
35) Dibromofluoromethane	7.914	113	264930	51.348	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.700%
50) Toluene-d8	10.338	98	994546	50.423	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.840%
62) 4-Bromofluorobenzene	12.620	95	302232	45.847	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100621\
Data File : VD070585.D
Acq On : 06 Oct 2021 11:19
Operator : VA/SY
Sample : VD1006SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1006SBL01

Quant Time: Oct 07 06:58:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

