

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100622\
 Data File : VD074465.D
 Acq On : 06 Oct 2022 14:45
 Operator : VA/SY
 Sample : N5014-10
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-10-0-2.0

Quant Time: Oct 07 01:26:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 01:54:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	12052	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	23179	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	23790	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	11823	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	10446	97.555	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	195.100%#
35) Dibromofluoromethane	7.799	113	10693	69.960	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	139.920%
50) Toluene-d8	10.269	98	26531	47.612	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.220%
62) 4-Bromofluorobenzene	12.569	95	10186	60.317	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	120.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100622\
Data File : VD074465.D
Acq On : 06 Oct 2022 14:45
Operator : VA/SY
Sample : N5014-10
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-10-0-2.0

Quant Time: Oct 07 01:26:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 01:54:40 2022
Response via : Initial Calibration

