

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100622\
 Data File : VD074471.D
 Acq On : 06 Oct 2022 17:32
 Operator : VA/SY
 Sample : N5014-08
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-09-0-2.0

Quant Time: Oct 07 01:27:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 01:54:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	133111	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	235672	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	206663	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	89980	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	48257	40.804	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	81.600%
35) Dibromofluoromethane	7.804	113	67053	43.147	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.300%
50) Toluene-d8	10.269	98	217777	38.788	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	77.580%
62) 4-Bromofluorobenzene	12.569	95	74250	41.246	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	82.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100622\
Data File : VD074471.D
Acq On : 06 Oct 2022 17:32
Operator : VA/SY
Sample : N5014-08
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-09-0-2.0

Quant Time: Oct 07 01:27:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 01:54:40 2022
Response via : Initial Calibration

