

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100622\
 Data File : VD074473.D
 Acq On : 06 Oct 2022 18:28
 Operator : VA/SY
 Sample : N5014-11
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-10-0-2.0-DUP

Quant Time: Oct 07 01:27:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 01:54:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	22551	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	44435	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	45686	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	21209	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	18466	92.165	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	184.320%#	
35) Dibromofluoromethane	7.804	113	17688	60.367	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	120.740%	
50) Toluene-d8	10.269	98	42977	40.513	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	81.020%	
62) 4-Bromofluorobenzene	12.569	95	17821	54.431	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	108.860%	
Target Compounds						
20) Methylene Chloride	4.799	84	3698	7.563	ug/l	Qvalue # 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100622\
Data File : VD074473.D
Acq On : 06 Oct 2022 18:28
Operator : VA/SY
Sample : N5014-11
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-10-0-2.0-DUP

Quant Time: Oct 07 01:27:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 01:54:40 2022
Response via : Initial Calibration

