



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100622\
Data File : VD074458.D
Acq On : 06 Oct 2022 10:31
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
Title : SW846 8260
Last Update : Thu Oct 06 01:54:40 2022



AutoFind: Scans 1866, 1867, 1868; Background Corrected with Scan 1858

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.4	9313	PASS
75	95	30	60	50.7	27171	PASS
95	95	100	100	100.0	53621	PASS
96	95	5	9	6.4	3424	PASS
173	174	0.00	2	1.1	548	PASS
174	95	50	100	91.0	48811	PASS
175	174	5	9	8.2	3986	PASS
176	174	95	101	96.9	47283	PASS
177	176	5	9	6.5	3083	PASS