



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100623\
Data File : VD077283.D
Acq On : 06 Oct 2023 10:47
Operator : JC/SY
Sample : BFB
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100623S.M
Title : SW846 8260
Last Update : Mon Oct 09 09:52:06 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1858

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.0	25888	PASS
75	95	30	60	50.5	68856	PASS
95	95	100	100	100.0	136379	PASS
96	95	5	9	6.5	8800	PASS
173	174	0.00	2	1.0	1167	PASS
174	95	50	100	89.7	122291	PASS
175	174	5	9	6.8	8289	PASS
176	174	95	101	96.5	117976	PASS
177	176	5	9	6.4	7602	PASS