

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100721\
 Data File : VD070605.D
 Acq On : 07 Oct 2021 09:24
 Operator : VA/SY
 Sample : VD1007SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1007SBL01

Quant Time: Oct 08 04:48:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	407957	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	788903	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	734363	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	295395	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	247780	51.61	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.22%
35) Dibromofluoromethane	7.914	113	262922	50.13	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.26%
50) Toluene-d8	10.337	98	1002110	49.98	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.96%
62) 4-Bromofluorobenzene	12.626	95	303388	45.27	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.54%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100721\
Data File : VD070605.D
Acq On : 07 Oct 2021 09:24
Operator : VA/SY
Sample : VD1007SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1007SBL01

Quant Time: Oct 08 04:48:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

