

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100721\
 Data File : VD070611.D
 Acq On : 07 Oct 2021 12:28
 Operator : VA/SY
 Sample : M4061-03
 Misc : 4.76G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-3

Quant Time: Oct 08 04:50:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	378635	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	739901	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	690114	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	277767	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	229368	51.47	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.94%	
35) Dibromofluoromethane	7.908	113	250595	50.94	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.88%	
50) Toluene-d8	10.337	98	945525	50.28	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.56%	
62) 4-Bromofluorobenzene	12.626	95	279892	44.53	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 89.06%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100721\
Data File : VD070611.D
Acq On : 07 Oct 2021 12:28
Operator : VA/SY
Sample : M4061-03
Misc : 4.76G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-3

Quant Time: Oct 08 04:50:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

