

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100721\
 Data File : VD070618.D
 Acq On : 07 Oct 2021 15:47
 Operator : VA/SY
 Sample : M4113-05
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-3

Quant Time: Oct 08 04:52:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	11744	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	19590	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	19884	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	8139	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	10880	78.720	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	157.440%
35) Dibromofluoromethane	7.908	113	7545	57.929	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.860%
50) Toluene-d8	10.332	98	24271	48.746	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.500%
62) 4-Bromofluorobenzene	12.620	95	6902	41.475	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	82.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100721\
Data File : VD070618.D
Acq On : 07 Oct 2021 15:47
Operator : VA/SY
Sample : M4113-05
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-3

Quant Time: Oct 08 04:52:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

