

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100721\
 Data File : VD070621.D
 Acq On : 07 Oct 2021 17:12
 Operator : VA/SY
 Sample : M4117-02
 Misc : 2.50G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-1(3-3.5)

Quant Time: Oct 08 04:53:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	360391	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	684913	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	654151	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	264146	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	216467	51.04	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.08%	
35) Dibromofluoromethane	7.914	113	230895	50.70	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.40%	
50) Toluene-d8	10.338	98	891045	51.19	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.38%	
62) 4-Bromofluorobenzene	12.620	95	263384	45.27	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 90.54%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100721\
Data File : VD070621.D
Acq On : 07 Oct 2021 17:12
Operator : VA/SY
Sample : M4117-02
Misc : 2.50G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-1(3-3.5)

Quant Time: Oct 08 04:53:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

