

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100722\
 Data File : VD074484.D
 Acq On : 07 Oct 2022 12:09
 Operator : VA/SY
 Sample : N5014-12RE
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-10-2.0-3.5RE

Quant Time: Oct 08 02:24:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 01:54:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	7075	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	12867	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	11835	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	5526	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	5181	82.422	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	164.840%#
35) Dibromofluoromethane	7.804	113	4689	55.265	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.520%
50) Toluene-d8	10.269	98	11214	36.686	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	73.380%
62) 4-Bromofluorobenzene	12.569	95	4339	44.644	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100722\
Data File : VD074484.D
Acq On : 07 Oct 2022 12:09
Operator : VA/SY
Sample : N5014-12RE
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-10-2.0-3.5RE

Quant Time: Oct 08 02:24:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 01:54:40 2022
Response via : Initial Calibration

