

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100722\
 Data File : VD074485.D
 Acq On : 07 Oct 2022 12:37
 Operator : VA/SY
 Sample : N5014-04RE
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-06-0-2.0RE

Quant Time: Oct 08 02:24:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 01:54:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	76910	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	145097	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	130661	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	54239	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	39068	57.174	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.340%
35) Dibromofluoromethane	7.804	113	43980	45.966	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.940%
50) Toluene-d8	10.269	98	125766	36.495	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	73.000%
62) 4-Bromofluorobenzene	12.569	95	44712	40.188	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100722\
Data File : VD074485.D
Acq On : 07 Oct 2022 12:37
Operator : VA/SY
Sample : N5014-04RE
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-06-0-2.0RE

Quant Time: Oct 08 02:24:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 01:54:40 2022
Response via : Initial Calibration

