

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100722\
 Data File : VD074487.D
 Acq On : 07 Oct 2022 13:33
 Operator : VA/SY
 Sample : N5014-02
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-05-0-2.0

Quant Time: Oct 08 02:24:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 01:54:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	116753	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	210072	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	182628	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	75424	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	50183	48.378	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.760%
35) Dibromofluoromethane	7.805	113	61574	44.450	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.900%
50) Toluene-d8	10.269	98	242122	47.930	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.860%
62) 4-Bromofluorobenzene	12.569	95	61958	38.162	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	76.320%
Target Compounds						
					Qvalue	
16) Acetone	4.022	43	1297	16.292	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100722\
Data File : VD074487.D
Acq On : 07 Oct 2022 13:33
Operator : VA/SY
Sample : N5014-02
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-05-0-2.0

Quant Time: Oct 08 02:24:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 01:54:40 2022
Response via : Initial Calibration

