

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100722\
 Data File : VD074488.D
 Acq On : 07 Oct 2022 14:00
 Operator : VA/SY
 Sample : N5014-03RE
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-05-2.0-3.0RE

Quant Time: Oct 08 02:24:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 01:54:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	2517	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	5369	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	5035	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	2404	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	2200	98.378	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 196.760%#	
35) Dibromofluoromethane	7.792	113	1927	54.429	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.860%	
50) Toluene-d8	10.269	98	6444	49.837	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 99.680%	
62) 4-Bromofluorobenzene	12.574	95	1875	46.485	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 92.960%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100722\
Data File : VD074488.D
Acq On : 07 Oct 2022 14:00
Operator : VA/SY
Sample : N5014-03RE
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-05-2.0-3.0RE

Quant Time: Oct 08 02:24:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 01:54:40 2022
Response via : Initial Calibration

