

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100722\
 Data File : VD074492.D
 Acq On : 07 Oct 2022 15:52
 Operator : VA/SY
 Sample : N5014-13
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-01-0-2.0

Quant Time: Oct 08 02:25:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 01:54:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	109564	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	197854	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	173756	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	71375	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	50678	52.061	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.120%
35) Dibromofluoromethane	7.804	113	60654	46.490	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.980%
50) Toluene-d8	10.269	98	236568	49.655	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.300%
62) 4-Bromofluorobenzene	12.569	95	60501	39.825	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	79.660%
Target Compounds						
16) Acetone	4.016	43	3520	24.511	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100722\
Data File : VD074492.D
Acq On : 07 Oct 2022 15:52
Operator : VA/SY
Sample : N5014-13
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-01-0-2.0

Quant Time: Oct 08 02:25:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 01:54:40 2022
Response via : Initial Calibration

