

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100722\
 Data File : VD074493.D
 Acq On : 07 Oct 2022 16:20
 Operator : VA/SY
 Sample : N5014-14RE
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-02-0-2.0RE

Quant Time: Oct 08 02:25:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 01:54:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	106009	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	188571	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	167046	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	70658	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	92098	97.784	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 195.560%#	
35) Dibromofluoromethane	7.804	113	111662	89.800	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 179.600%#	
50) Toluene-d8	10.269	98	461492	99.735	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 199.460%#	
62) 4-Bromofluorobenzene	12.569	95	120640	91.026	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 182.060%#	
Target Compounds						Qvalue
16) Acetone	4.022	43	1488	17.433	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100722\
Data File : VD074493.D
Acq On : 07 Oct 2022 16:20
Operator : VA/SY
Sample : N5014-14RE
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-02-0-2.0RE

Quant Time: Oct 08 02:25:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 01:54:40 2022
Response via : Initial Calibration

