

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100722\
 Data File : VD074480.D
 Acq On : 07 Oct 2022 09:00
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 08 02:22:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 01:54:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.441	0.441	0.0	90	0.00
3 P	Chloromethane	0.609	0.549	9.9	91	0.00
4 C	Vinyl Chloride	0.528	0.515	2.5#	90	0.00
5 T	Bromomethane	0.309	0.333	-7.8	111	0.00
6 T	Chloroethane	0.348	0.335	3.7	92	0.00
7 T	Trichlorofluoromethane	0.865	0.848	2.0	92	0.00
8 T	Diethyl Ether	0.268	0.268	0.0	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.559	0.562	-0.5	96	0.00
10 T	Methyl Iodide	0.603	0.625	-3.6	88	0.00
11 T	Tert butyl alcohol	0.069	0.034	50.7#	78	0.00
12 CM	1,1-Dichloroethene	0.589	0.565	4.1#	91	0.00
13 T	Acrolein	0.029	0.021	27.6#	65	0.00
14 T	Allyl chloride	0.675	0.637	5.6	89	0.00
15 T	Acrylonitrile	0.106	0.106	0.0	92	0.00
16 T	Acetone	0.098	0.097	1.0	78	0.00
17 T	Carbon Disulfide	1.860	1.795	3.5	92	0.00
18 T	Methyl Acetate	0.270	0.241	10.7	87	0.00
19 T	Methyl tert-butyl Ether	1.161	1.141	1.7	92	0.00
20 T	Methylene Chloride	0.830	0.635	23.5	93	0.00
21 T	trans-1,2-Dichloroethene	0.652	0.632	3.1	94	0.00
22 T	Diisopropyl ether	1.375	1.328	3.4	91	0.00
23 T	Vinyl Acetate	0.632	0.646	-2.2	89	0.00
24 P	1,1-Dichloroethane	0.998	0.975	2.3	94	0.00
25 T	2-Butanone	0.127	0.124	2.4	84	0.00
26 T	2,2-Dichloropropane	0.881	0.862	2.2	97	0.00
27 T	cis-1,2-Dichloroethene	0.689	0.682	1.0	99	0.00
28 T	Bromochloromethane	0.295	0.296	-0.3	95	0.00
29 T	Tetrahydrofuran	0.071	0.070	1.4	93	0.00
30 C	Chloroform	1.031	0.997	3.3#	98	0.00
31 T	Cyclohexane	0.892	0.840	5.8	95	0.00
32 T	1,1,1-Trichloroethane	0.911	0.895	1.8	99	0.00
33 S	1,2-Dichloroethane-d4	0.444	0.458	-3.2	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.330	0.305	7.6	95	0.00
36 T	1,1-Dichloropropene	0.463	0.474	-2.4	98	0.00
37 T	Ethyl Acetate	0.152	0.159	-4.6	93	0.00
38 T	Carbon Tetrachloride	0.410	0.447	-9.0	101	0.00
39 T	Methylcyclohexane	0.567	0.562	0.9	90	0.00
40 TM	Benzene	1.344	1.425	-6.0	100	0.00
41 T	Methacrylonitrile	0.083	0.088	-6.0	114	0.00
42 TM	1,2-Dichloroethane	0.318	0.322	-1.3	95	0.00
43 T	Isopropyl Acetate	0.290	0.290	0.0	92	0.00
44 TM	Trichloroethene	0.378	0.373	1.3	95	0.00
45 C	1,2-Dichloropropane	0.312	0.296	5.1#	90	0.00
46 T	Dibromomethane	0.173	0.171	1.2	90	0.00
47 T	Bromodichloromethane	0.432	0.419	3.0	92	0.00
48 T	Methyl methacrylate	0.135	0.130	3.7	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100722\
 Data File : VD074480.D
 Acq On : 07 Oct 2022 09:00
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 08 02:22:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 01:54:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	87	0.00
50 S	Toluene-d8	1.123	1.220	-8.6	92	0.00
51 T	4-Methyl-2-Pentanone	0.144	0.143	0.7	87	0.00
52 CM	Toluene	0.850	0.860	-1.2#	92	0.00
53 T	t-1,3-Dichloropropene	0.394	0.388	1.5	89	0.00
54 T	cis-1,3-Dichloropropene	0.485	0.479	1.2	91	0.00
55 T	1,1,2-Trichloroethane	0.243	0.232	4.5	91	0.00
56 T	Ethyl methacrylate	0.273	0.278	-1.8	89	0.00
57 T	1,3-Dichloropropane	0.392	0.379	3.3	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.054	0.063	-16.7	100	0.00
59 T	2-Hexanone	0.099	0.101	-2.0	83	0.00
60 T	Dibromochloromethane	0.283	0.277	2.1	92	0.00
61 T	1,2-Dibromoethane	0.225	0.227	-0.9	93	0.00
62 S	4-Bromofluorobenzene	0.421	0.341	19.0	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.347	0.343	1.2	91	0.00
65 PM	Chlorobenzene	0.999	1.038	-3.9	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.336	0.351	-4.5	93	0.00
67 C	Ethyl Benzene	1.801	1.905	-5.8#	93	0.00
68 T	m/p-Xylenes	0.708	0.750	-5.9	91	0.00
69 T	o-Xylene	0.650	0.686	-5.5	93	0.00
70 T	Styrene	1.114	1.188	-6.6	91	0.00
71 P	Bromoform	0.177	0.189	-6.8	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	3.677	3.909	-6.3	92	0.00
74 T	N-amyl acetate	0.629	0.640	-1.7	86	0.00
75 P	1,1,2,2-Tetrachloroethane	0.611	0.625	-2.3	89	0.00
76 T	1,2,3-Trichloropropane	0.384	0.374	2.6	94	0.00
77 T	Bromobenzene	0.836	0.866	-3.6	93	0.00
78 T	n-propylbenzene	4.492	4.690	-4.4	90	0.00
79 T	2-Chlorotoluene	2.393	2.487	-3.9	90	0.00
80 T	1,3,5-Trimethylbenzene	3.038	3.213	-5.8	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.172	0.188	-9.3	88	0.00
82 T	4-Chlorotoluene	2.443	2.604	-6.6	92	0.00
83 T	tert-Butylbenzene	2.614	2.735	-4.6	90	0.00
84 T	1,2,4-Trimethylbenzene	2.985	3.175	-6.4	91	0.00
85 T	sec-Butylbenzene	4.027	4.241	-5.3	91	0.00
86 T	p-Isopropyltoluene	3.322	3.557	-7.1	92	0.00
87 T	1,3-Dichlorobenzene	1.692	1.775	-4.9	91	0.00
88 T	1,4-Dichlorobenzene	1.684	1.772	-5.2	93	0.00
89 T	n-Butylbenzene	3.111	3.340	-7.4	91	0.00
90 T	Hexachloroethane	0.589	0.620	-5.3	92	0.00
91 T	1,2-Dichlorobenzene	1.439	1.510	-4.9	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.090	0.088	2.2	91	0.00
93 T	1,2,4-Trichlorobenzene	0.934	0.946	-1.3	90	0.00
94 T	Hexachlorobutadiene	0.487	0.476	2.3	88	0.00
95 T	Naphthalene	1.782	1.852	-3.9	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100722\
Data File : VD074480.D
Acq On : 07 Oct 2022 09:00
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Oct 08 02:22:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 01:54:40 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.809	0.806	0.4	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6