

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD100820\
 Data File : VD067201.D
 Acq On : 08 Oct 2020 18:16
 Operator : VA/SY
 Sample : L4313-02
 Misc : 6.30G/5.00ml/MSVOA D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-1

Quant Time: Oct 09 03:40:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 15:19:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	378769	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	607354	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	558703	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	267278	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	183206	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
35) Dibromofluoromethane	7.92	113	191296	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.22%	
50) Toluene-d8	10.34	98	711647	49.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.24%	
62) 4-Bromofluorobenzene	12.64	95	234193	47.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.22%	
Target Compounds						
20) Methylene Chloride	4.97	84	61946	9.830	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD100820\
Data File : VD067201.D
Acq On : 08 Oct 2020 18:16
Operator : VA/SY
Sample : L4313-02
Misc : 6.30G/5.00ml/MSVOA D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-1

Quant Time: Oct 09 03:40:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 15:19:17 2020
Response via : Initial Calibration

