


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD100820\
Data File : VD067182.D
Acq On : 08 Oct 2020 09:25
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
Title : SW846 8260
Last Update : Thu Oct 01 15:19:17 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	39304	PASS
75	95	30	60	50.9	118837	PASS
95	95	100	100	100.0	233536	PASS
96	95	5	9	7.1	16509	PASS
173	174	0.00	2	0.7	1421	PASS
174	95	50	100	87.9	205376	PASS
175	174	5	9	7.9	16129	PASS
176	174	95	101	97.3	199829	PASS
177	176	5	9	6.8	13511	PASS