

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100821\
 Data File : VD070629.D
 Acq On : 08 Oct 2021 10:38
 Operator : VA/SY
 Sample : VD1008SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1008SBL01

Quant Time: Oct 08 18:10:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	399629	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	759551	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	703244	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	290035	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	239080	50.835	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.660%
35) Dibromofluoromethane	7.914	113	254680	50.432	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.860%
50) Toluene-d8	10.332	98	970130	50.252	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.500%
62) 4-Bromofluorobenzene	12.620	95	290516	45.025	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100821\
Data File : VD070629.D
Acq On : 08 Oct 2021 10:38
Operator : VA/SY
Sample : VD1008SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1008SBL01

Quant Time: Oct 08 18:10:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

