

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100821\
 Data File : VD070632.D
 Acq On : 08 Oct 2021 12:02
 Operator : VA/SY
 Sample : M4061-02
 Misc : 5.12G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-2

Quant Time: Oct 11 02:21:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	417803	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	791380	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	725878	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	299987	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	249704	50.784	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.560%
35) Dibromofluoromethane	7.909	113	260263	49.465	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.940%
50) Toluene-d8	10.338	98	988625	49.151	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.300%
62) 4-Bromofluorobenzene	12.626	95	303952	45.213	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100821\
Data File : VD070632.D
Acq On : 08 Oct 2021 12:02
Operator : VA/SY
Sample : M4061-02
Misc : 5.12G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-2

Quant Time: Oct 11 02:21:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

