

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100821\
 Data File : VD070636.D
 Acq On : 08 Oct 2021 13:56
 Operator : VA/SY
 Sample : M4113-05RE
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-3RE

Quant Time: Oct 11 02:22:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	12203	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.861	114	20232	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	19431	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	7147	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	10187	70.934	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	141.860%
35) Dibromofluoromethane	7.908	113	6580	48.917	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.840%
50) Toluene-d8	10.338	98	23486	45.672	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.340%
62) 4-Bromofluorobenzene	12.620	95	7297	42.457	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100821\
Data File : VD070636.D
Acq On : 08 Oct 2021 13:56
Operator : VA/SY
Sample : M4113-05RE
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-3RE

Quant Time: Oct 11 02:22:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

