

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100821\
 Data File : VD070639.D
 Acq On : 08 Oct 2021 15:22
 Operator : VA/SY
 Sample : M4113-03
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-2

Quant Time: Oct 11 02:23:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	378938	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	719146	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	668555	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	269839	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	212640	47.682	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.360%
35) Dibromofluoromethane	7.914	113	236070	49.374	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.740%
50) Toluene-d8	10.338	98	925452	50.631	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.260%
62) 4-Bromofluorobenzene	12.620	95	269594	44.130	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	88.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100821\
Data File : VD070639.D
Acq On : 08 Oct 2021 15:22
Operator : VA/SY
Sample : M4113-03
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-2

Quant Time: Oct 11 02:23:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

