

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100821\
 Data File : VD070646.D
 Acq On : 08 Oct 2021 18:41
 Operator : VA/SY
 Sample : M3990-11
 Misc : 6.29G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-836-K1-SO-9-10-092921

Quant Time: Oct 11 02:24:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	378299	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	719001	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	675332	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	278277	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	235670	52.935	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.880%	
35) Dibromofluoromethane	7.914	113	244602	51.168	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.340%	
50) Toluene-d8	10.338	98	921775	50.440	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.880%	
62) 4-Bromofluorobenzene	12.620	95	289481	47.395	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 94.800%	
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	8962	11.854	ug/l	93
17) Carbon Disulfide	4.408	76	51594	3.183	ug/l #	93
27) cis-1,2-Dichloroethene	7.208	96	16640	3.033	ug/l	88
44) Trichloroethene	9.114	130	13462	2.426	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD100821\
Data File : VD070646.D
Acq On : 08 Oct 2021 18:41
Operator : VA/SY
Sample : M3990-11
Misc : 6.29G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-836-K1-SO-9-10-092921

Quant Time: Oct 11 02:24:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

