

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD100918\
 Data File : VD060164.D
 Acq On : 9 Oct 2018 21:26
 Operator : VA/AP
 Sample : J5326-29
 Misc : 6.05g/5ml/MSVOA D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FL-PIT-8

Quant Time: Oct 10 02:12:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 06:56:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.37	168	1239828	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	7.41	114	1707291	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.27	117	1386424	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.36	152	724932	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.75	65	622680	66.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	132.52%	
35) Dibromofluoromethane	6.29	113	790465	60.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.94%	
50) Toluene-d8	9.43	98	2173950	58.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.60%	
62) 4-Bromofluorobenzene	12.41	95	819208	59.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.36%	
Target Compounds						
16) Acetone	3.12	43	40739	23.889	ug/l	95
20) Methylene Chloride	3.67	84	76042	1.493	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD100918\
Data File : VD060164.D
Acq On : 9 Oct 2018 21:26
Operator : VA/AP
Sample : J5326-29
Misc : 6.05g/5ml/MSVOA D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FL-PIT-8

Quant Time: Oct 10 02:12:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D100118S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 06:56:35 2018
Response via : Initial Calibration

