

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101022\
 Data File : VD074505.D
 Acq On : 10 Oct 2022 16:28
 Operator : VA/SY
 Sample : N5058-05
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BP-VPB-RW5-GW-823-825

Quant Time: Oct 11 08:21:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 01:54:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	103652	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	182051	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	152897	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	54790	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	44234	48.033	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.060%
35) Dibromofluoromethane	7.804	113	56978	47.463	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.920%
50) Toluene-d8	10.269	98	225599	51.397	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.800%
62) 4-Bromofluorobenzene	12.569	95	55758	39.900	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	79.800%
Target Compounds						Qvalue
16) Acetone	4.022	43	1663	18.217	ug/l	# 37

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101022\
Data File : VD074505.D
Acq On : 10 Oct 2022 16:28
Operator : VA/SY
Sample : N5058-05
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BP-VPB-RW5-GW-823-825

Quant Time: Oct 11 08:21:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100522S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 01:54:40 2022
Response via : Initial Calibration

