

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101121\
 Data File : VD070653.D
 Acq On : 11 Oct 2021 11:55
 Operator : VA/SY
 Sample : VD1011SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1011SBL01

Quant Time: Oct 12 05:38:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	441082	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	823927	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	764724	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	319696	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	249233	48.01	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.02%
35) Dibromofluoromethane	7.908	113	263382	48.08	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.16%
50) Toluene-d8	10.337	98	1030495	49.21	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.42%
62) 4-Bromofluorobenzene	12.626	95	326321	46.62	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.24%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101121\
Data File : VD070653.D
Acq On : 11 Oct 2021 11:55
Operator : VA/SY
Sample : VD1011SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1011SBL01

Quant Time: Oct 12 05:38:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

